

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071425\
 Data File : VX046981.D
 Acq On : 14 Jul 2025 17:53
 Operator : JC/MD
 Sample : Q2579-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-A2-13-G

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/15/2025
 Supervised By :Semsettin Yesilyurt 07/15/2025

Quant Time: Jul 15 01:25:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	354879	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	603011	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	549695	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	285450	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	245018	52.262	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 104.520%			
35) Dibromofluoromethane	5.568	113	739m	0.181	ug/l	0.17
Spiked Amount 50.000	Range 75 - 124		Recovery = 0.360%#			
50) Toluene-d8	8.653	98	699230	48.417	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 96.840%			
62) 4-Bromofluorobenzene	11.079	95	285220	51.965	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 103.920%			
Target Compounds						
16) Acetone	2.386	43	101530	69.719	ug/l	93
20) Methylene Chloride	2.800	84	22186	5.152	ug/l	96
95) Naphthalene	13.774	128	1724711	106.357	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071425\
Data File : VX046981.D
Acq On : 14 Jul 2025 17:53
Operator : JC/MD
Sample : Q2579-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-A2-13-G

Quant Time: Jul 15 01:25:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/15/2025
Supervised By :Semsettin Yesilyurt 07/15/2025

