

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071425\
 Data File : VX046982.D
 Acq On : 14 Jul 2025 18:14
 Operator : JC/MD
 Sample : Q2579-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-A2-14-G

Quant Time: Jul 15 01:25:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	353132	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	601642	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	548369	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	283025	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	247347	53.019	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.040%
35) Dibromofluoromethane	5.580	113	587	0.144	ug/l	0.18
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.280%#
50) Toluene-d8	8.653	98	698421	48.471	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.940%
62) 4-Bromofluorobenzene	11.079	95	283567	51.781	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.560%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	100306	69.219	ug/l	95
20) Methylene Chloride	2.806	84	23062	5.382	ug/l	95
95) Naphthalene	13.774	128	1539105	95.725	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071425\
Data File : VX046982.D
Acq On : 14 Jul 2025 18:14
Operator : JC/MD
Sample : Q2579-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-A2-14-G

Quant Time: Jul 15 01:25:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

