

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071425\
 Data File : VX046984.D
 Acq On : 14 Jul 2025 18:56
 Operator : JC/MD
 Sample : Q2553-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB

Quant Time: Jul 14 20:01:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	356948	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	602998	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	548783	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	281291	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	248196	52.633	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.260%
35) Dibromofluoromethane	5.397	113	199290	48.688	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.380%
50) Toluene-d8	8.653	98	691698	47.896	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.800%
62) 4-Bromofluorobenzene	11.079	95	284303	51.799	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.600%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	3233	2.207	ug/l #	68
17) Carbon Disulfide	2.526	76	9009	0.888	ug/l	96
20) Methylene Chloride	2.800	84	2370	0.547	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071425\
Data File : VX046984.D
Acq On : 14 Jul 2025 18:56
Operator : JC/MD
Sample : Q2553-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB

Quant Time: Jul 14 20:01:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

