

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071519\
 Data File : VX010872.D
 Acq On : 15 Jul 2019 10:25
 Operator : JC/SP
 Sample : K3788-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-7-12-19

Quant Time: Jul 16 02:15:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	242531	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	380751	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	350172	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	162556	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	127390	56.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.04%	
35) Dibromofluoromethane	5.49	113	119917	51.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.96%	
50) Toluene-d8	8.71	98	456883	51.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.52%	
62) 4-Bromofluorobenzene	11.13	95	159476	49.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.12%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071519\
Data File : VX010872.D
Acq On : 15 Jul 2019 10:25
Operator : JC/SP
Sample : K3788-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-7-12-19

Quant Time: Jul 16 02:15:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

