

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071519\
 Data File : VX010875.D
 Acq On : 15 Jul 2019 11:36
 Operator : JC/SP
 Sample : K3788-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 2056

Quant Time: Jul 16 02:27:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	236955	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	372487	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	343068	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	162137	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	124960	56.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.50%	
35) Dibromofluoromethane	5.50	113	118636	52.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.12%	
50) Toluene-d8	8.71	98	445111	51.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.10%	
62) 4-Bromofluorobenzene	11.13	95	156798	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
Target Compounds						
16) Acetone	2.44	43	6375	5.725	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071519\
Data File : VX010875.D
Acq On : 15 Jul 2019 11:36
Operator : JC/SP
Sample : K3788-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
2056

Quant Time: Jul 16 02:27:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

