

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071519\
 Data File : VX010884.D
 Acq On : 15 Jul 2019 15:04
 Operator : JC/SP
 Sample : K3817-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW533-2126

Quant Time: Jul 16 03:25:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	226674	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	353567	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	323928	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	146041	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	119211	56.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.18%	
35) Dibromofluoromethane	5.49	113	111634	51.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.20%	
50) Toluene-d8	8.71	98	423592	51.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.36%	
62) 4-Bromofluorobenzene	11.13	95	146179	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
Target Compounds						
16) Acetone	2.44	43	3197	3.002	ug/l	94
24) 1,1-Dichloroethane	3.70	63	3447	1.023	ug/l	93
85) sec-Butylbenzene	11.94	105	3473	0.361	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071519\
Data File : VX010884.D
Acq On : 15 Jul 2019 15:04
Operator : JC/SP
Sample : K3817-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW533-2126

Quant Time: Jul 16 03:25:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

