

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071519\
 Data File : VX010886.D
 Acq On : 15 Jul 2019 15:51
 Operator : JC/SP
 Sample : K3786-01DL 4X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL-0581DL

Manual Integrations
 APPROVED

MMDadoda
 7/16/2019 10:44:28 AM

Quant Time: Jul 16 03:30:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	230487	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	360166	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	332977	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	162269	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	120162	55.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.20%	
35) Dibromofluoromethane	5.50	113	112281	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
50) Toluene-d8	8.71	98	431022	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
62) 4-Bromofluorobenzene	11.13	95	156787	51.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.02%	
Target Compounds						
					Qvalue	
16) Acetone	2.44	43	1290	1.191	ug/l	90
31) Cyclohexane	5.57	56	55856	17.894	ug/l	91
39) Methylcyclohexane	7.46	83	57372	15.250	ug/l	95
40) Benzene	6.14	78	8972	1.006	ug/l	96
67) Ethyl Benzene	10.25	91	302874	26.539	ug/l	98
68) m/p-Xylenes	10.35	106	295738	66.878	ug/l	96
69) o-Xylene	10.69	106	5744	1.321	ug/l	97
73) Isopropylbenzene	11.02	105	30992	2.817	ug/l	98
78) n-propylbenzene	11.36	91	39200	3.243	ug/l	97
80) 1,3,5-Trimethylbenzene	11.51	105	84554	9.324	ug/l	100
84) 1,2,4-Trimethylbenzene	11.80	105	320544	35.184	ug/l	99
85) sec-Butylbenzene	11.94	105	5860m	0.548	ug/l	
86) p-Isopropyltoluene	12.06	119	10696	1.082	ug/l	96
95) Naphthalene	13.83	128	72277	6.365	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071519\
Data File : VX010886.D
Acq On : 15 Jul 2019 15:51
Operator : JC/SP
Sample : K3786-01DL 4X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL-0581DL

Manual Integrations
APPROVED

MMDadoda
7/16/2019 10:44:28 AM

Quant Time: Jul 16 03:30:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

