

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071519\
 Data File : VX010891.D
 Acq On : 15 Jul 2019 17:47
 Operator : JC/SP
 Sample : K3786-07DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL-0597DL

Manual Integrations
 APPROVED

MMDadoda
 7/16/2019 10:47:44 AM

Quant Time: Jul 16 03:44:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	229302	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	358393	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	334788	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	162635	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	119061	55.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.76%	
35) Dibromofluoromethane	5.50	113	110801	50.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.06%	
50) Toluene-d8	8.71	98	432993	51.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.22%	
62) 4-Bromofluorobenzene	11.13	95	153422	50.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.32%	
Target Compounds						
					Qvalue	
16) Acetone	2.45	43	2200	2.042	ug/l #	88
31) Cyclohexane	5.58	56	28055	9.034	ug/l	96
39) Methylcyclohexane	7.46	83	39519	10.556	ug/l	100
67) Ethyl Benzene	10.25	91	100207	8.733	ug/l	97
68) m/p-Xylenes	10.35	106	223433	50.254	ug/l	100
69) o-Xylene	10.69	106	62171	14.219	ug/l	100
73) Isopropylbenzene	11.02	105	31635	2.869	ug/l	96
78) n-propylbenzene	11.36	91	37557	3.100	ug/l	100
80) 1,3,5-Trimethylbenzene	11.51	105	107850	11.867	ug/l	99
84) 1,2,4-Trimethylbenzene	11.80	105	318084	34.835	ug/l	99
85) sec-Butylbenzene	11.94	105	6183m	0.577	ug/l	
86) p-Isopropyltoluene	12.06	119	6918	0.698	ug/l	90
95) Naphthalene	13.83	128	50345	4.423	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071519\
Data File : VX010891.D
Acq On : 15 Jul 2019 17:47
Operator : JC/SP
Sample : K3786-07DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL-0597DL

Manual Integrations
APPROVED

MMDadoda
7/16/2019 10:47:44 AM

Quant Time: Jul 16 03:44:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

