

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071519\
 Data File : VX010892.D
 Acq On : 15 Jul 2019 18:11
 Operator : JC/SP
 Sample : K3791-01DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL-0598DL

Manual Integrations
 APPROVED

MMDadoda
 7/16/2019 10:47:48 AM

Quant Time: Jul 16 03:47:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	229466	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	363441	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	334253	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	159182	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	120162	55.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.70%	
35) Dibromofluoromethane	5.49	113	112125	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
50) Toluene-d8	8.71	98	435087	51.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.28%	
62) 4-Bromofluorobenzene	11.13	95	154098	50.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.34%	
Target Compounds						
					Qvalue	
31) Cyclohexane	5.58	56	14694	4.728	ug/l	93
39) Methylcyclohexane	7.46	83	10605	2.793	ug/l	92
67) Ethyl Benzene	10.25	91	97716	8.530	ug/l	97
68) m/p-Xylenes	10.35	106	149368	33.649	ug/l	97
69) o-Xylene	10.70	106	20682	4.738	ug/l	94
73) Isopropylbenzene	11.02	105	15290	1.417	ug/l	98
78) n-propylbenzene	11.36	91	8678	0.732	ug/l	96
80) 1,3,5-Trimethylbenzene	11.51	105	12111	1.361	ug/l	95
84) 1,2,4-Trimethylbenzene	11.80	105	76778	8.591	ug/l	99
85) sec-Butylbenzene	11.94	105	4253	0.406	ug/l #	75
86) p-Isopropyltoluene	12.06	119	2578m	0.266	ug/l	
95) Naphthalene	13.83	128	39531	3.549	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071519\
Data File : VX010892.D
Acq On : 15 Jul 2019 18:11
Operator : JC/SP
Sample : K3791-01DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL-0598DL

Manual Integrations
APPROVED

MMDadoda
7/16/2019 10:47:48 AM

Quant Time: Jul 16 03:47:20 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

