

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071519\
 Data File : VX010893.D
 Acq On : 15 Jul 2019 18:34
 Operator : JC/SP
 Sample : K3791-08DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL- 0594DL

Quant Time: Jul 16 03:53:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	235698	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	366105	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	339737	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	166664	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	120649	54.59	ug/l	0.00
Spiked Amount			Recovery	=	109.18%	
35) Dibromofluoromethane	5.50	113	111979	49.99	ug/l	0.00
Spiked Amount			Recovery	=	99.98%	
50) Toluene-d8	8.71	98	440291	51.38	ug/l	0.00
Spiked Amount			Recovery	=	102.76%	
62) 4-Bromofluorobenzene	11.13	95	157724	50.98	ug/l	0.00
Spiked Amount			Recovery	=	101.96%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
31) Cyclohexane	5.58	56	69093	21.645	ug/l	94
39) Methylcyclohexane	7.46	83	56499	14.774	ug/l	91
67) Ethyl Benzene	10.25	91	321803	27.636	ug/l	100
68) m/p-Xylenes	10.36	106	155403	34.443	ug/l	96
69) o-Xylene	10.69	106	22911	5.164	ug/l	97
73) Isopropylbenzene	11.02	105	31277	2.768	ug/l	98
78) n-propylbenzene	11.36	91	49562	3.992	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	62150	6.673	ug/l	97
84) 1,2,4-Trimethylbenzene	11.80	105	226789	24.237	ug/l	99
85) sec-Butylbenzene	11.94	105	9656	0.880	ug/l	81
86) p-Isopropyltoluene	12.06	119	7938	0.782	ug/l	94
89) n-Butylbenzene	12.38	91	10724	1.243	ug/l #	35
95) Naphthalene	13.83	128	59246	5.080	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071519\
Data File : VX010893.D
Acq On : 15 Jul 2019 18:34
Operator : JC/SP
Sample : K3791-08DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL- 0594DL

Quant Time: Jul 16 03:53:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

