

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071519\
 Data File : VX010882.D
 Acq On : 15 Jul 2019 14:17
 Operator : JC/SP
 Sample : K3817-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-20190709

Quant Time: Jul 16 03:21:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	237496	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	373161	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	347169	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	161416	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	124822	56.06	ug/l	0.00
Spiked Amount			Recovery	=	112.12%	
35) Dibromofluoromethane	5.50	113	116152	50.87	ug/l	0.00
Spiked Amount			Recovery	=	101.74%	
50) Toluene-d8	8.71	98	446046	51.07	ug/l	0.00
Spiked Amount			Recovery	=	102.14%	
62) 4-Bromofluorobenzene	11.13	95	158035	50.11	ug/l	0.00
Spiked Amount			Recovery	=	100.22%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071519\
Data File : VX010882.D
Acq On : 15 Jul 2019 14:17
Operator : JC/SP
Sample : K3817-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-20190709

Quant Time: Jul 16 03:21:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.00 min

Lab File: VX010882.D

Acq: 15 Jul 2019 14:17

Instrument :

MSVOA_X

ClientSampleId :


TB01-20190709

Tot Ion:168 Resp: 237496

Ion Ratio Lower Upper




168 100

99 48.4 35.5 53.3



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 56.055 ug/l

RT: 6.06 min Scan# 805

Delta R.T. 0.00 min

Lab File: VX010882.D


Acq: 15 Jul 2019 14:17

Tgt Ion: 65 Resp: 124822

Ion Ratio Lower Upper



65 100


67 53.7 0.0 110.2



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. 0.00 min
 Lab File: VX010882.D
 Acq: 15 Jul 2019 14:17




Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-20190709


Tot Ion	Ratio	Lower	Upper
114	100		
63	18.0	0.0	33.8
88	14.3	0.0	27.6



#35
 Dibromofluoromethane
 Concen: 50.874 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. 0.00 min
 Lab File: VX010882.D
 Acq: 15 Jul 2019 14:17

Tgt Ion	Ratio	Lower	Upper
113	100		
111	101.8	81.2	121.8
192	22.0	18.6	27.8









#50
Toluene-d8
Concen: 51.065 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. 0.00 min
Lab File: VX010882.D
Acq: 15 Jul 2019 14:17




Instrument :
MSVOA_X
ClientSampleId :
TB01-20190709


Tot Ion: 98 Resp: 446046
Ion Ratio Lower Upper
98 100
100 63.6 51.6 77.4



#62
4-Bromofluorobenzene
Concen: 50.114 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. 0.00 min
Lab File: VX010882.D
Acq: 15 Jul 2019 14:17

Tgt Ion: 95 Resp: 158035
Ion Ratio Lower Upper
95 100
174 90.8 0.0 187.6
176 87.5 0.0 184.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. 0.00 min
Lab File: VX010882.D
Acq: 15 Jul 2019 14:17

Instrument :
MSVOA_X
ClientSampleId :
TB01-20190709

Tot Ion:117 Resp: 347169
Ion Ratio Lower Upper
117 100
82 51.2 41.2 61.8
119 31.8 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX010882.D
Acq: 15 Jul 2019 14:17

Tgt Ion:152 Resp: 161416
Ion Ratio Lower Upper
152 100
115 54.9 38.6 115.7
150 156.9 0.0 347.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

