

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
 Data File : VX023294.D
 Acq On : 15 Jul 2021 10:20
 Operator : JC/MD
 Sample : VX0715MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0715MBL01

Quant Time: Jul 16 03:50:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	207868	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	333489	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	281320	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102305	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	138963	49.741	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.480%
35) Dibromofluoromethane	5.391	113	107696	50.009	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.020%
50) Toluene-d8	8.653	98	394792	49.145	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.300%
62) 4-Bromofluorobenzene	11.085	95	128779	41.525	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.040%
Target Compounds						
20) Methylene Chloride	2.794	84	2331	0.980	ug/l	Qvalue # 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
Data File : VX023294.D
Acq On : 15 Jul 2021 10:20
Operator : JC/MD
Sample : VX0715MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0715MBL01

Quant Time: Jul 16 03:50:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

