

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
 Data File : VX023295.D
 Acq On : 15 Jul 2021 13:25
 Operator : JC/MD
 Sample : VX0715WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0715WBL01

Quant Time: Jul 16 03:50:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	239671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	387963	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	340481	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137268	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	161612	50.172	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117		Recovery	= 100.340%	
35) Dibromofluoromethane	5.397	113	122399	48.856	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124		Recovery	= 97.720%	
50) Toluene-d8	8.653	98	453473	48.524	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112		Recovery	= 97.040%	
62) 4-Bromofluorobenzene	11.085	95	158031	43.802	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123		Recovery	= 87.600%	
Target Compounds						
16) Acetone	2.386	43	3596	2.822	ug/l	98
20) Methylene Chloride	2.794	84	1332	0.485	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
Data File : VX023295.D
Acq On : 15 Jul 2021 13:25
Operator : JC/MD
Sample : VX0715WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0715WBL01

Quant Time: Jul 16 03:50:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

