

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
 Data File : VX023298.D
 Acq On : 15 Jul 2021 14:52
 Operator : JC/MD
 Sample : M3058-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 210714005-05-TRIP-BLANK

Quant Time: Jul 16 03:51:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	220629	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	359606	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	317131	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135435	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	148563	50.101	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.200%
35) Dibromofluoromethane	5.397	113	113578	48.910	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.820%
50) Toluene-d8	8.653	98	424143	48.964	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.920%
62) 4-Bromofluorobenzene	11.085	95	151911	45.426	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.860%
Target Compounds						Qvalue
16) Acetone	2.386	43	3068	2.615	ug/l	97
20) Methylene Chloride	2.794	84	1047	0.415	ug/l	94
30) Chloroform	5.099	83	3673	0.835	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
Data File : VX023298.D
Acq On : 15 Jul 2021 14:52
Operator : JC/MD
Sample : M3058-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
210714005-05-TRIP-BLANK

Quant Time: Jul 16 03:51:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

