

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
 Data File : VX023299.D
 Acq On : 15 Jul 2021 15:15
 Operator : JC/MD
 Sample : M3027-03 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NWB-1639

Quant Time: Jul 16 03:51:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	222537	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	361702	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	324075	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	142620	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	153527	51.331	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 102.660%			
35) Dibromofluoromethane	5.397	113	114886	49.187	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.380%			
50) Toluene-d8	8.653	98	426349	48.934	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 97.860%			
62) 4-Bromofluorobenzene	11.085	95	158956	47.257	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 94.520%			
Target Compounds						
16) Acetone	2.386	43	22297	18.843	ug/l	94
95) Naphthalene	13.780	128	2651	1.518	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
Data File : VX023299.D
Acq On : 15 Jul 2021 15:15
Operator : JC/MD
Sample : M3027-03 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NWB-1639

Quant Time: Jul 16 03:51:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

