

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
 Data File : VX023303.D
 Acq On : 15 Jul 2021 16:48
 Operator : JC/MD
 Sample : M3026-06RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ABN-JG-2RE

Quant Time: Jul 16 03:52:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	220985	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	353780	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	318463	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132132	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	152218	51.251	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.500%
35) Dibromofluoromethane	5.397	113	55377	24.240	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	48.480%#
50) Toluene-d8	8.653	98	420487	49.342	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.680%
62) 4-Bromofluorobenzene	11.085	95	151424	46.026	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.060%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	130206	110.810	ug/l	89
20) Methylene Chloride	2.794	84	19231	7.602	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
Data File : VX023303.D
Acq On : 15 Jul 2021 16:48
Operator : JC/MD
Sample : M3026-06RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ABN-JG-2RE

Quant Time: Jul 16 03:52:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

