

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
 Data File : VX023304.D
 Acq On : 15 Jul 2021 17:12
 Operator : JC/MD
 Sample : M3026-09RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ABN-JG-3RE

Quant Time: Jul 16 03:53:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	216370	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	354491	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	309735	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	128748	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	152611	52.480	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.960%	
35) Dibromofluoromethane	5.397	113	52736	23.037	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 46.080%#	
50) Toluene-d8	8.653	98	415691	48.681	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.360%	
62) 4-Bromofluorobenzene	11.085	95	147898	44.864	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.720%	
Target Compounds						Qvalue
16) Acetone	2.386	43	132393	115.075	ug/l	91
20) Methylene Chloride	2.794	84	19245	7.769	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
Data File : VX023304.D
Acq On : 15 Jul 2021 17:12
Operator : JC/MD
Sample : M3026-09RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ABN-JG-3RE

Quant Time: Jul 16 03:53:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

