

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
 Data File : VX023305.D
 Acq On : 15 Jul 2021 17:35
 Operator : JC/MD
 Sample : M3026-12RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ABN-JG-4RE

Quant Time: Jul 16 03:53:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	229663	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	374630	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	329889	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138100	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	161346	52.272	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 104.540%			
35) Dibromofluoromethane	5.391	113	56128	23.201	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 46.400%#			
50) Toluene-d8	8.653	98	441074	48.877	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 97.760%			
62) 4-Bromofluorobenzene	11.085	95	156104	44.808	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 89.620%			
Target Compounds						Qvalue
16) Acetone	2.385	43	137297	112.430	ug/l	90
20) Methylene Chloride	2.794	84	20389	7.755	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
Data File : VX023305.D
Acq On : 15 Jul 2021 17:35
Operator : JC/MD
Sample : M3026-12RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ABN-JG-4RE

Quant Time: Jul 16 03:53:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

