

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
 Data File : VX023307.D
 Acq On : 15 Jul 2021 18:22
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 16 03:54:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	249258	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	402237	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	363010	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	159985	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	167633	50.039	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.080%	
35) Dibromofluoromethane	5.397	113	130916	50.401	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.800%	
50) Toluene-d8	8.653	98	477001	49.230	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.460%	
62) 4-Bromofluorobenzene	11.085	95	177162	47.362	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.720%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.971	59	1776	2.520	ug/l #	91
16) Acetone	2.392	43	2773	2.092	ug/l #	82
20) Methylene Chloride	2.794	84	2908	1.019	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
Data File : VX023307.D
Acq On : 15 Jul 2021 18:22
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jul 16 03:54:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

