

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
 Data File : VX023309.D
 Acq On : 15 Jul 2021 19:10
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 16 03:54:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	231375	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	376227	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	331053	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	140663	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	153346	49.313	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.620%
35) Dibromofluoromethane	5.397	113	120761	49.706	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.420%
50) Toluene-d8	8.653	98	440475	48.603	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.200%
62) 4-Bromofluorobenzene	11.085	95	158796	45.387	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.780%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.977	59	2126	3.250	ug/l #	87
16) Acetone	2.385	43	2495	2.028	ug/l	91
20) Methylene Chloride	2.800	84	1903	0.718	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071521\
Data File : VX023309.D
Acq On : 15 Jul 2021 19:10
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jul 16 03:54:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

