

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071522\
 Data File : VX030096.D
 Acq On : 15 Jul 2022 17:16
 Operator : JC/MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDC0050EC

Quant Time: Jul 16 06:03:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	178421	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	312842	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	278827	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138104	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	117767	54.228	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.460%	
35) Dibromofluoromethane	5.391	113	99858	49.965	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.940%	
50) Toluene-d8	8.653	98	385735	51.319	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.640%	
62) 4-Bromofluorobenzene	11.085	95	141292	54.168	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.340%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	106489	52.598	ug/l	98
3) Chloromethane	1.288	50	95060	51.129	ug/l	98
4) Vinyl Chloride	1.374	62	105569	51.121	ug/l	100
5) Bromomethane	1.605	94	40907	46.631	ug/l	95
6) Chloroethane	1.678	64	57214	49.199	ug/l	98
7) Trichlorofluoromethane	1.880	101	141089	52.902	ug/l	98
8) Diethyl Ether	2.136	74	53014	48.049	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.325	101	97070	53.315	ug/l	100
10) Methyl Iodide	2.447	142	90136	50.404	ug/l	97
11) Tert butyl alcohol	2.983	59	124188	221.569	ug/l	97
12) 1,1-Dichloroethene	2.312	96	96183	52.911	ug/l	97
13) Acrolein	2.239	56	19022	170.402	ug/l	99
14) Allyl chloride	2.660	41	173453	54.601	ug/l	98
15) Acrylonitrile	3.068	53	343349	269.540	ug/l	100
16) Acetone	2.386	43	274273	268.022	ug/l	96
17) Carbon Disulfide	2.508	76	238224	46.807	ug/l	99
18) Methyl Acetate	2.709	43	165593	53.192	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	311422	54.197	ug/l	96
20) Methylene Chloride	2.788	84	108441	52.509	ug/l	98
21) trans-1,2-Dichloroethene	3.093	96	100997	53.302	ug/l	97
22) Diisopropyl ether	3.763	45	349746	56.245	ug/l	90
23) Vinyl Acetate	3.721	43	1531367	290.910	ug/l	98
24) 1,1-Dichloroethane	3.611	63	184450	53.530	ug/l	98
25) 2-Butanone	4.562	43	475282	272.655	ug/l	96
26) 2,2-Dichloropropane	4.477	77	127463	52.660	ug/l	100
27) cis-1,2-Dichloroethene	4.489	96	120358	51.871	ug/l	99
28) Bromochloromethane	4.903	49	76229	52.586	ug/l	99
29) Tetrahydrofuran	5.007	42	320821	271.193	ug/l	99
30) Chloroform	5.092	83	187869	54.711	ug/l	99
31) Cyclohexane	5.470	56	172905	54.921	ug/l	96
32) 1,1,1-Trichloroethane	5.391	97	151272	53.900	ug/l	98
36) 1,1-Dichloropropene	5.696	75	137181	53.140	ug/l	99
37) Ethyl Acetate	4.715	43	178790	53.899	ug/l	98
38) Carbon Tetrachloride	5.678	117	129052	53.608	ug/l	94
39) Methylcyclohexane	7.379	83	181238	55.382	ug/l	96
40) Benzene	6.037	78	437769	53.878	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071522\
 Data File : VX030096.D
 Acq On : 15 Jul 2022 17:16
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Jul 16 06:03:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	96669	54.123	ug/l	98
42) 1,2-Dichloroethane	6.092	62	142601	56.630	ug/l	100
43) Isopropyl Acetate	6.342	43	270632	54.256	ug/l	99
44) Trichloroethene	7.123	130	118268	49.812	ug/l	97
45) 1,2-Dichloropropane	7.434	63	113640	53.808	ug/l	98
46) Dibromomethane	7.580	93	77097	53.039	ug/l	99
47) Bromodichloromethane	7.824	83	142344	53.967	ug/l	98
48) Methyl methacrylate	7.696	41	132209	57.444	ug/l	98
49) 1,4-Dioxane	7.677	88	63227	1057.384	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	917787	278.052	ug/l	99
52) Toluene	8.720	92	269945	52.319	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	149975	54.741	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	167112	53.244	ug/l	96
55) 1,1,2-Trichloroethane	9.153	97	112802	53.016	ug/l	98
56) Ethyl methacrylate	9.116	69	177103	55.362	ug/l	96
57) 1,3-Dichloropropane	9.311	76	182757	54.071	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	484063	278.668	ug/l	99
59) 2-Hexanone	9.433	43	717618	283.506	ug/l	99
60) Dibromochloromethane	9.525	129	107529	51.874	ug/l	97
61) 1,2-Dibromoethane	9.610	107	117555	53.086	ug/l	98
64) Tetrachloroethene	9.275	164	117915	52.682	ug/l	98
65) Chlorobenzene	10.079	112	282205	50.827	ug/l	97
66) 1,1,1,2-Tetrachloroethane	10.165	131	99028	52.317	ug/l	99
67) Ethyl Benzene	10.195	91	516299	54.866	ug/l	99
68) m/p-Xylenes	10.305	106	393750	106.684	ug/l	97
69) o-Xylene	10.640	106	191702	51.874	ug/l	96
70) Styrene	10.659	104	329525	53.723	ug/l	98
71) Bromoform	10.799	173	73074	48.051	ug/l #	98
73) Isopropylbenzene	10.963	105	506991	52.723	ug/l	100
74) N-amyl acetate	10.848	43	231546	54.823	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	171514	49.437	ug/l	99
76) 1,2,3-Trichloropropane	11.244	75	202055	63.633	ug/l	86
77) Bromobenzene	11.201	156	115566	48.033	ug/l	97
78) n-propylbenzene	11.305	91	616254	54.019	ug/l	99
79) 2-Chlorotoluene	11.366	91	358393	52.450	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	430106	53.081	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	56539	51.387	ug/l	96
82) 4-Chlorotoluene	11.457	91	407861	53.274	ug/l	100
83) tert-Butylbenzene	11.719	119	397423	51.145	ug/l	95
84) 1,2,4-Trimethylbenzene	11.756	105	430931	53.768	ug/l	99
85) sec-Butylbenzene	11.890	105	553175	53.849	ug/l	100
86) p-Isopropyltoluene	12.012	119	447565	54.308	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	229421	51.608	ug/l	100
88) 1,4-Dichlorobenzene	12.042	146	230189	50.458	ug/l	99
89) n-Butylbenzene	12.335	91	424667	57.512	ug/l	100
90) Hexachloroethane	12.542	117	71270	50.588	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	222887	49.351	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	40415	51.018	ug/l	94
93) 1,2,4-Trichlorobenzene	13.591	180	139445	50.268	ug/l	98
94) Hexachlorobutadiene	13.725	225	54556	49.259	ug/l	98
95) Naphthalene	13.780	128	555044	50.408	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	140755	49.867	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071522\
Data File : VX030096.D
Acq On : 15 Jul 2022 17:16
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Quant Time: Jul 16 06:03:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071522\
 Data File : VX030096.D
 Acq On : 15 Jul 2022 17:16
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Jul 16 06:03:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:168 Resp: 178421
Ion Ratio Lower Upper
168 100
99 56.8 44.0 66.0



#2
Dichlorodifluoromethane
Concen: 52.598 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion: 85 Resp: 106489
Ion Ratio Lower Upper
85 100
87 31.9 16.5 49.5









#3
Chloromethane
Concen: 51.129 ug/l
RT: 1.288 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 50 Resp: 95060
Ion Ratio Lower Upper
50 100
52 31.8 26.3 39.5



#4
Vinyl Chloride
Concen: 51.121 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion: 62 Resp: 105569
Ion Ratio Lower Upper
62 100
64 31.5 25.2 37.8









#5
Bromomethane
Concen: 46.631 ug/l
RT: 1.605 min Scan# 81
Delta R.T. -0.006 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 94 Resp: 40907
Ion Ratio Lower Upper
94 100
96 91.3 76.9 115.3



#6
Chloroethane
Concen: 49.199 ug/l
RT: 1.678 min Scan# 97
Delta R.T. -0.007 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion: 64 Resp: 57214
Ion Ratio Lower Upper
64 100
66 32.9 27.2 40.8









#7
 Trichlorofluoromethane
 Concen: 52.902 ug/l
 RT: 1.880 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: VX030096.D
 Acq: 15 Jul 2022 17:16




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:101 Resp: 141089
 Ion Ratio Lower Upper
 101 100
 103 63.4 49.7 74.5



#8
 Diethyl Ether
 Concen: 48.049 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX030096.D
 Acq: 15 Jul 2022 17:16

Tgt Ion: 74 Resp: 53014
 Ion Ratio Lower Upper
 74 100
 45 103.7 51.5 154.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 53.315 ug/l
RT: 2.325 min Scan# 204
Delta R.T. -0.006 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 97070
Ion Ratio Lower Upper
101 100
85 42.6 34.0 51.0
151 74.1 59.4 89.0



#10
Methyl Iodide
Concen: 50.404 ug/l
RT: 2.447 min Scan# 223
Delta R.T. -0.006 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion:142 Resp: 90136
Ion Ratio Lower Upper
142 100
127 40.0 30.2 45.2
141 13.8 11.3 16.9









#11
Tert butyl alcohol
Concen: 221.569 ug/l
RT: 2.983 min Scan# 31
Delta R.T. -0.012 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 59 Resp: 124188
Ion Ratio Lower Upper
59 100
57 10.4 7.4 11.2



#12
1,1-Dichloroethene
Concen: 52.911 ug/l
RT: 2.312 min Scan# 201
Delta R.T. -0.007 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion: 96 Resp: 96183
Ion Ratio Lower Upper
96 100
61 151.4 117.4 176.0
98 66.3 52.2 78.4









#13
Acrolein
Concen: 170.402 ug/l
RT: 2.239 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 56 Resp: 19022
Ion Ratio Lower Upper
56 100
55 71.8 56.6 84.8



#14
Allyl chloride
Concen: 54.601 ug/l
RT: 2.660 min Scan# 258
Delta R.T. -0.006 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion: 41 Resp: 173453
Ion Ratio Lower Upper
41 100
39 63.0 48.9 73.3
76 32.9 26.9 40.3









#15
Acrylonitrile
Concen: 269.540 ug/l
RT: 3.068 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 53 Resp: 343349
Ion Ratio Lower Upper
53 100
52 82.0 65.4 98.0
51 34.9 27.7 41.5



#16
Acetone
Concen: 268.022 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion: 43 Resp: 274273
Ion Ratio Lower Upper
43 100
58 32.4 27.7 41.5









#17
Carbon Disulfide
Concen: 46.807 ug/l
RT: 2.508 min Scan# 21
Delta R.T. -0.006 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 238224
Ion Ratio Lower Upper
76 100
78 9.2 7.8 11.6



#18
Methyl Acetate
Concen: 53.192 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion: 43 Resp: 165593
Ion Ratio Lower Upper
43 100
74 23.8 19.4 29.2









#19
Methyl tert-butyl Ether
Concen: 54.197 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 73 Resp: 311422
Ion Ratio Lower Upper
73 100
57 21.8 19.0 28.6



#20
Methylene Chloride
Concen: 52.509 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion: 84 Resp: 108441
Ion Ratio Lower Upper
84 100
49 119.1 93.8 140.6
51 38.3 29.4 44.0
86 64.2 51.0 76.4









#21
trans-1,2-Dichloroethene
Concen: 53.302 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 96 Resp: 100997
Ion Ratio Lower Upper
96 100
61 128.2 105.6 158.4
98 63.6 50.1 75.1



#22
Diisopropyl ether
Concen: 56.245 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion: 45 Resp: 349746
Ion Ratio Lower Upper
45 100
43 69.9 65.4 98.0
87 26.7 22.3 33.5
59 10.8 9.0 13.4









#23
 Vinyl Acetate
 Concen: 290.910 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: VX030096.D
 Acq: 15 Jul 2022 17:16




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 43 Resp: 1531367
 Ion Ratio Lower Upper
 43 100
 86 9.9 8.5 12.7



#24
 1,1-Dichloroethane
 Concen: 53.530 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.000 min
 Lab File: VX030096.D
 Acq: 15 Jul 2022 17:16

Tgt Ion: 63 Resp: 184450
 Ion Ratio Lower Upper
 63 100
 98 7.7 3.5 10.5
 100 5.0 2.3 6.9









#25
2-Butanone
Concen: 272.655 ug/l
RT: 4.562 min Scan# 51
Delta R.T. -0.006 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 475282
Ion Ratio Lower Upper
43 100
72 25.2 21.7 32.5



#26
2,2-Dichloropropane
Concen: 52.660 ug/l
RT: 4.477 min Scan# 556
Delta R.T. -0.006 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion: 77 Resp: 127463
Ion Ratio Lower Upper
77 100
97 25.9 13.0 39.0





#27

cis-1,2-Dichloroethene

Concen: 51.871 ug/l

RT: 4.489 min Scan# 51

Delta R.T. -0.006 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 96 Resp: 120358

Ion Ratio Lower Upper

96 100

61 136.6 0.0 270.0

98 64.8 0.0 127.8



#28

Bromochloromethane

Concen: 52.586 ug/l

RT: 4.903 min Scan# 626

Delta R.T. 0.000 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16




Tgt Ion: 49 Resp: 76229


Ion Ratio Lower Upper

49 100

129 1.2 0.0 4.4

130 79.6 62.8 94.2





#29

Tetrahydrofuran

Concen: 271.193 ug/l

RT: 5.007 min Scan# 644

Delta R.T. -0.006 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC




Tgt Ion: 42 Resp: 320821

Ion Ratio Lower Upper

42 100

72 45.0 36.5 54.7

71 41.8 33.9 50.9



#30

Chloroform

Concen: 54.711 ug/l

RT: 5.092 min Scan# 657

Delta R.T. -0.013 min

Lab File: VX030096.D




Acq: 15 Jul 2022 17:16


Tgt Ion: 83 Resp: 187869

Ion Ratio Lower Upper

83 100

85 63.7 51.5 77.3









#31
Cyclohexane
Concen: 54.921 ug/l
RT: 5.470 min Scan# 719
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 172905
Ion Ratio Lower Upper
56 100
69 30.0 25.4 38.0
84 82.1 68.9 103.3



#32
1,1,1-Trichloroethane
Concen: 53.900 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion: 97 Resp: 151272
Ion Ratio Lower Upper
97 100
99 66.1 51.4 77.0
61 45.7 35.8 53.8





#33

1,2-Dichloroethane-d4

Concen: 54.228 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 65 Resp: 117767

Ion Ratio Lower Upper

65 100

67 55.2 0.0 109.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.000 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16




Tgt Ion: 114 Resp: 312842


Ion Ratio Lower Upper

114 100

63 19.8 0.0 38.4

88 15.9 0.0 31.8





#35

Dibromofluoromethane

Concen: 49.965 ug/l

RT: 5.391 min Scan# 706

Delta R.T. 0.000 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:113 Resp: 99858

Ion Ratio Lower Upper

113 100

111 104.2 83.2 124.8

192 18.5 14.8 22.2



#36

1,1-Dichloropropene

Concen: 53.140 ug/l

RT: 5.696 min Scan# 756

Delta R.T. 0.000 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16




Tgt Ion: 75 Resp: 137181


Ion Ratio Lower Upper

75 100

110 38.2 19.3 57.9

77 31.1 24.4 36.6





#37

Ethyl Acetate

Concen: 53.899 ug/l

RT: 4.715 min Scan# 595

Delta R.T. -0.006 min

Lab File: VX030096.D


Acq: 15 Jul 2022 17:16

Instrument :




MSVOA_X

ClientSampleId :

VSTDCCC050EC



Tgt Ion:	43	Resp:	178790
Ion	Ratio	Lower	Upper
43	100		
61	14.2	11.7	17.5
70	10.5	9.3	13.9



#38

Carbon Tetrachloride

Concen: 53.608 ug/l




RT: 5.678 min Scan# 753


Delta R.T. -0.006 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16

Tgt Ion:	117	Resp:	129052
Ion	Ratio	Lower	Upper
117	100		
119	99.5	74.1	111.1
121	30.2	25.8	38.8





#39

Methylcyclohexane

Concen: 55.382 ug/l

RT: 7.379 min Scan# 1

Delta R.T. -0.006 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83 Resp: 181238

Ion Ratio Lower Upper

83 100

55 76.7 57.7 86.5

98 46.4 36.5 54.7



#40

Benzene

Concen: 53.878 ug/l

RT: 6.037 min Scan# 812

Delta R.T. -0.007 min

Lab File: VX030096.D




Acq: 15 Jul 2022 17:16


Tgt Ion: 78 Resp: 437769

Ion Ratio Lower Upper

78 100

77 24.1 18.0 27.0









#41
Methacrylonitrile
Concen: 54.123 ug/l
RT: 4.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	41	Resp:	96669
Ion	Ratio	Lower	Upper
41	100		
39	56.2	42.9	64.3
67	74.8	60.1	90.1
52	30.0	22.6	34.0



#42
1,2-Dichloroethane
Concen: 56.630 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion:	62	Resp:	142601
Ion	Ratio	Lower	Upper
62	100		
98	9.9	0.0	20.2









#43
Isopropyl Acetate
Concen: 54.256 ug/l
RT: 6.342 min Scan# 80
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 270632
Ion Ratio Lower Upper
43 100
61 20.0 16.5 24.7
87 13.8 11.1 16.7



#44
Trichloroethene
Concen: 49.812 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.006 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion:130 Resp: 118268
Ion Ratio Lower Upper
130 100
95 100.2 0.0 194.8





#45

1,2-Dichloropropane

Concen: 53.808 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 63 Resp: 113640

Ion Ratio Lower Upper

63 100

65 32.4 25.2 37.8



#46

Dibromomethane

Concen: 53.039 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16




Tgt Ion: 93 Resp: 77097


Ion Ratio Lower Upper

93 100

95 85.1 67.4 101.0

174 93.1 75.6 113.4









#47
Bromodichloromethane
Concen: 53.967 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 83 Resp: 142344
Ion Ratio Lower Upper
83 100
85 65.2 53.5 80.3
127 8.6 6.6 10.0



#48
Methyl methacrylate
Concen: 57.444 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion: 41 Resp: 132209
Ion Ratio Lower Upper
41 100
69 83.4 68.8 103.2
39 57.1 44.6 66.8





#49

1,4-Dioxane

Concen: 1057.384 ug/l

RT: 7.677 min Scan# 1081

Delta R.T. -0.006 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 88 Resp: 63227

Ion Ratio Lower Upper

88 100

43 30.1 23.8 35.8

58 69.3 53.8 80.6



#50

Toluene-d8

Concen: 51.319 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX030096.D




Acq: 15 Jul 2022 17:16

Tgt Ion: 98 Resp: 385735

Ion Ratio Lower Upper

98 100

100 65.4 52.3 78.5





#51

4-Methyl-2-Pentanone

Concen: 278.052 ug/l

RT: 8.574 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 43 Resp: 917787

Ion Ratio Lower Upper

43 100

58 38.9 31.4 47.2



#52

Toluene

Concen: 52.319 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. -0.000 min

Lab File: VX030096.D




Acq: 15 Jul 2022 17:16


Tgt Ion: 92 Resp: 269945

Ion Ratio Lower Upper

92 100

91 174.9 139.2 208.8





#53

t-1,3-Dichloropropene

Concen: 54.741 ug/l

RT: 8.982 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 75 Resp: 149975

Ion Ratio Lower Upper

75 100

77 32.5 26.3 39.5



#54

cis-1,3-Dichloropropene

Concen: 53.244 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.006 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16




Tgt Ion: 75 Resp: 167112


Ion Ratio Lower Upper

75 100

77 32.6 25.0 37.4

39 50.5 37.5 56.3





#55

1,1,2-Trichloroethane

Concen: 53.016 ug/l

RT: 9.153 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 97 Resp: 112802





Ion Ratio Lower Upper

97 100

83 84.8 68.9 103.3

85 56.3 44.0 66.0

99 62.5 48.3 72.5



#56

Ethyl methacrylate

Concen: 55.362 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16




Tgt Ion: 69 Resp: 177103


Ion Ratio Lower Upper

69 100

41 67.8 52.3 78.5

39 37.9 28.2 42.4









#57
1,3-Dichloropropane
Concen: 54.071 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 182757
Ion Ratio Lower Upper
76 100
78 32.7 25.9 38.9



#58
2-Chloroethyl Vinyl ether
Concen: 278.668 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion: 63 Resp: 484063
Ion Ratio Lower Upper
63 100
106 27.4 22.2 33.2





#59

2-Hexanone

Concen: 283.506 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 43 Resp: 717618

Ion Ratio Lower Upper

43 100

58 54.5 27.6 82.8



#60

Dibromochloromethane

Concen: 51.874 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VX030096.D




Acq: 15 Jul 2022 17:16


Tgt Ion: 129 Resp: 107529

Ion Ratio Lower Upper

129 100

127 79.5 38.5 115.5









#61
1,2-Dibromoethane
Concen: 53.086 ug/l
RT: 9.610 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:107 Resp: 117555
Ion Ratio Lower Upper
107 100
109 93.4 76.6 114.8



#62
4-Bromofluorobenzene
Concen: 54.168 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.006 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion: 95 Resp: 141292
Ion Ratio Lower Upper
95 100
174 73.7 0.0 149.0
176 71.0 0.0 146.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:117 Resp: 278827
Ion Ratio Lower Upper
117 100
82 53.3 42.7 64.1
119 31.9 25.4 38.2



#64
Tetrachloroethene
Concen: 52.682 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion:164 Resp: 117915
Ion Ratio Lower Upper
164 100
166 121.9 98.2 147.4
129 96.1 80.8 121.2
131 96.1 75.5 113.3








#65
Chlorobenzene
Concen: 50.827 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:112 Resp: 282205
Ion Ratio Lower Upper
112 100
114 32.6 24.6 36.8



#66
1,1,1,2-Tetrachloroethane
Concen: 52.317 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion:131 Resp: 99028
Ion Ratio Lower Upper
131 100
133 95.8 46.9 140.8
119 68.1 34.4 103.1









#67
Ethyl Benzene
Concen: 54.866 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 516299
Ion Ratio Lower Upper
91 100
106 31.4 25.5 38.3



#68
m/p-Xylenes
Concen: 106.684 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion:106 Resp: 393750
Ion Ratio Lower Upper
106 100
91 202.7 158.2 237.4









#69
o-Xylene
Concen: 51.874 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.006 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:106 Resp: 191702
Ion Ratio Lower Upper
106 100
91 213.5 103.8 311.6



#70
Styrene
Concen: 53.723 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion:104 Resp: 329525
Ion Ratio Lower Upper
104 100
78 49.1 37.8 56.8
103 55.9 44.5 66.7









#71
Bromoform
Concen: 48.051 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.006 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:173 Resp: 73074
Ion Ratio Lower Upper
173 100
175 48.8 23.8 71.2
254 0.2 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion:152 Resp: 138104
Ion Ratio Lower Upper
152 100
115 83.9 40.5 121.5
150 176.6 0.0 347.0









#73
Isopropylbenzene
Concen: 52.723 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:105 Resp: 506991
Ion Ratio Lower Upper
105 100
120 26.8 13.5 40.4



#74
N-ethyl acetate
Concen: 54.823 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion: 43 Resp: 231546
Ion Ratio Lower Upper
43 100
70 40.4 32.3 48.5
55 22.7 18.4 27.6
61 23.1 19.2 28.8









#75
 1,1,2,2-Tetrachloroethane
 Concen: 49.437 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. 0.000 min
 Lab File: VX030096.D
 Acq: 15 Jul 2022 17:16




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tgt Ion: 83 Resp: 171514
 Ion Ratio Lower Upper
 83 100
 131 9.4 4.5 13.7
 85 64.3 32.5 97.5



#76
 1,2,3-Trichloropropane
 Concen: 63.633 ug/l
 RT: 11.244 min Scan# 1666
 Delta R.T. -0.000 min
 Lab File: VX030096.D
 Acq: 15 Jul 2022 17:16

Tgt Ion: 75 Resp: 202055
 Ion Ratio Lower Upper
 75 100
 77 31.1 19.8 59.3





#77

Bromobenzene

Concen: 48.033 ug/l

RT: 11.201 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:156 Resp: 115566

Ion Ratio Lower Upper

156 100

77 171.4 82.7 247.9

158 96.3 47.5 142.5



#78

n-propylbenzene

Concen: 54.019 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min

Lab File: VX030096.D




Acq: 15 Jul 2022 17:16


Tgt Ion: 91 Resp: 616254

Ion Ratio Lower Upper

91 100

120 22.7 11.6 34.8









#79
2-Chlorotoluene
Concen: 52.450 ug/l
RT: 11.366 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 358393
Ion Ratio Lower Upper
91 100
126 33.9 17.4 52.3



#80
1,3,5-Trimethylbenzene
Concen: 53.081 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion: 105 Resp: 430106
Ion Ratio Lower Upper
105 100
120 48.6 24.4 73.2








#81
trans-1,4-Dichloro-2-butene
Concen: 51.387 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 75 Resp: 56539
Ion Ratio Lower Upper
75 100
53 98.4 74.2 111.4
89 42.9 34.8 52.2



#82
4-Chlorotoluene
Concen: 53.274 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion: 91 Resp: 407861
Ion Ratio Lower Upper
91 100
126 30.6 15.3 45.9









#83
tert-Butylbenzene
Concen: 51.145 ug/l
RT: 11.719 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:119 Resp: 397423
Ion Ratio Lower Upper
119 100
91 61.0 28.3 84.9
134 23.4 11.2 33.5



#84
1,2,4-Trimethylbenzene
Concen: 53.768 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX030096.D
Acq: 15 Jul 2022 17:16

Tgt Ion:105 Resp: 430931
Ion Ratio Lower Upper
105 100
120 45.4 22.9 68.8









#85
 sec-Butylbenzene
 Concen: 53.849 ug/l
 RT: 11.890 min Scan# 11772
 Delta R.T. 0.000 min
 Lab File: VX030096.D
 Acq: 15 Jul 2022 17:16




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:105 Resp: 553175
 Ion Ratio Lower Upper
 105 100
 134 20.0 10.0 30.0



#86
 p-Isopropyltoluene
 Concen: 54.308 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX030096.D
 Acq: 15 Jul 2022 17:16

Tgt Ion:119 Resp: 447565
 Ion Ratio Lower Upper
 119 100
 134 26.9 13.9 41.6
 91 24.3 11.8 35.3









#87
 1,3-Dichlorobenzene
 Concen: 51.608 ug/l
 RT: 11.969 min Scan# 1785
 Delta R.T. 0.000 min
 Lab File: VX030096.D
 Acq: 15 Jul 2022 17:16




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:146 Resp: 229421
 Ion Ratio Lower Upper
 146 100
 111 40.4 20.3 60.8
 148 63.5 31.8 95.3



#88
 1,4-Dichlorobenzene
 Concen: 50.458 ug/l
 RT: 12.042 min Scan# 1797
 Delta R.T. 0.000 min
 Lab File: VX030096.D
 Acq: 15 Jul 2022 17:16

Tgt Ion:146 Resp: 230189
 Ion Ratio Lower Upper
 146 100
 111 40.1 20.3 60.9
 148 63.9 32.3 96.8





#89

n-Butylbenzene

Concen: 57.512 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 91 Resp: 424667

Ion Ratio Lower Upper

91 100

92 52.1 26.2 78.5

134 25.1 12.6 37.8



#90

Hexachloroethane

Concen: 50.588 ug/l

RT: 12.542 min Scan# 1879

Delta R.T. 0.000 min

Lab File: VX030096.D




Acq: 15 Jul 2022 17:16

Tgt Ion: 117 Resp: 71270

Ion Ratio Lower Upper

117 100

201 66.2 34.2 102.5









#91
 1,2-Dichlorobenzene
 Concen: 49.351 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. 0.000 min
 Lab File: VX030096.D
 Acq: 15 Jul 2022 17:16




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:146 Resp: 222887
 Ion Ratio Lower Upper
 146 100
 111 43.3 20.7 62.1
 148 64.4 32.0 96.2



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 51.018 ug/l
 RT: 12.945 min Scan# 1945
 Delta R.T. -0.000 min
 Lab File: VX030096.D
 Acq: 15 Jul 2022 17:16

Tgt Ion: 75 Resp: 40415
 Ion Ratio Lower Upper
 75 100
 155 85.6 45.0 134.8
 157 106.5 56.9 170.7





#93

1,2,4-Trichlorobenzene

Concen: 50.268 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. -0.000 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:180 Resp: 139445

Ion Ratio Lower Upper

180 100

182 93.9 48.2 144.6

145 31.8 15.9 47.6



#94

Hexachlorobutadiene

Concen: 49.259 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16




Tgt Ion:225 Resp: 54556


Ion Ratio Lower Upper

225 100

223 64.0 30.6 91.6

227 63.7 31.9 95.9





#95

Naphthalene

Concen: 50.408 ug/l

RT: 13.780 min Scan# 2082

Delta R.T. -0.000 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:128 Resp: 555044

Ion Ratio Lower Upper

128 100

127 12.6 10.2 15.2

129 10.8 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 49.867 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. -0.000 min

Lab File: VX030096.D

Acq: 15 Jul 2022 17:16

Tgt Ion:180 Resp: 140755

Ion Ratio Lower Upper

180 100

182 94.7 47.0 141.0

145 34.5 17.0 50.9

