

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071522\
 Data File : VX030088.D
 Acq On : 15 Jul 2022 13:53
 Operator : JC/MD
 Sample : N3735-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-071422

Quant Time: Jul 16 06:01:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	164901	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	294374	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240664	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91387	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	98375	49.012	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.020%
35) Dibromofluoromethane	5.391	113	87798	46.687	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.380%
50) Toluene-d8	8.653	98	334662	47.317	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.640%
62) 4-Bromofluorobenzene	11.085	95	102814	41.889	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.780%
Target Compounds						
16) Acetone	2.386	43	2698	2.853	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071522\
Data File : VX030088.D
Acq On : 15 Jul 2022 13:53
Operator : JC/MD
Sample : N3735-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-071422

Quant Time: Jul 16 06:01:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

