

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071522\
 Data File : VX030091.D
 Acq On : 15 Jul 2022 15:04
 Operator : JC/MD
 Sample : N3740-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jul 16 06:02:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	164670	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	288243	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	251337	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99159	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	98436	49.112	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.220%
35) Dibromofluoromethane	5.391	113	86367	46.903	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.800%
50) Toluene-d8	8.653	98	337245	48.697	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.400%
62) 4-Bromofluorobenzene	11.085	95	105684	43.974	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.940%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071522\
Data File : VX030091.D
Acq On : 15 Jul 2022 15:04
Operator : JC/MD
Sample : N3740-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jul 16 06:02:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

