

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071524\
 Data File : VX042316.D
 Acq On : 15 Jul 2024 10:18
 Operator : JC/MD
 Sample : VX0715MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0715MBL01

Quant Time: Jul 16 05:37:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	100655	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	153531	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127899	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51223	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59111	50.928	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.860%	
35) Dibromofluoromethane	5.385	113	56924	52.254	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.500%	
50) Toluene-d8	8.647	98	158245	41.349	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 82.700%#	
62) 4-Bromofluorobenzene	11.079	95	51365	36.410	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 72.820%#	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071524\
Data File : VX042316.D
Acq On : 15 Jul 2024 10:18
Operator : JC/MD
Sample : VX0715MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0715MBL01

Quant Time: Jul 16 05:37:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

