

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071524\
 Data File : VX042317.D
 Acq On : 15 Jul 2024 10:42
 Operator : JC/MD
 Sample : VX0715WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0715WBL01

Quant Time: Jul 16 05:37:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	136463	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	224494	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85575	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	89653	56.973	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 113.940%	
35) Dibromofluoromethane	5.385	113	86765	54.470	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.940%	
50) Toluene-d8	8.646	98	242341	43.306	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 86.620%#	
62) 4-Bromofluorobenzene	11.079	95	85939	41.661	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 83.320%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071524\
Data File : VX042317.D
Acq On : 15 Jul 2024 10:42
Operator : JC/MD
Sample : VX0715WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0715WBL01

Quant Time: Jul 16 05:37:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

