

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071524\
 Data File : VX042323.D
 Acq On : 15 Jul 2024 13:37
 Operator : JC/MD
 Sample : P3206-03
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 16777

Quant Time: Jul 16 05:39:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	154622	50.000	ug/l	#-0.01
34) 1,4-Difluorobenzene	6.757	114	237818	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	223606	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117364	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	84626	47.463	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.920%
35) Dibromofluoromethane	5.385	113	74266	44.011	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.020%
50) Toluene-d8	8.646	98	285096	48.092	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.180%
62) 4-Bromofluorobenzene	11.079	95	106840	48.892	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.780%
Target Compounds						
					Qvalue	
5) Bromomethane	1.593	94	827	1.131	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	7313	1.226	ug/l	95
95) Naphthalene	13.780	128	8681	1.764	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071524\
Data File : VX042323.D
Acq On : 15 Jul 2024 13:37
Operator : JC/MD
Sample : P3206-03
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
16777

Quant Time: Jul 16 05:39:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

