

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071524\
 Data File : VX042324.D
 Acq On : 15 Jul 2024 14:01
 Operator : JC/MD
 Sample : P3224-01 20X
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 380AVEC-21

Quant Time: Jul 16 05:39:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	138923	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	216786	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	187916	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85451	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85706	53.501	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.000%	
35) Dibromofluoromethane	5.391	113	81532	53.005	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.000%	
50) Toluene-d8	8.653	98	238071	44.056	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 88.120%#	
62) 4-Bromofluorobenzene	11.079	95	83931	42.134	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 84.260%	
Target Compounds						
					Qvalue	
40) Benzene	6.044	78	8199	1.488	ug/l	94
52) Toluene	8.720	92	10027	2.858	ug/l	98
68) m/p-Xylenes	10.305	106	7672	3.103	ug/l	91
69) o-Xylene	10.640	106	5387	2.201	ug/l	93
80) 1,3,5-Trimethylbenzene	11.451	105	5694	1.316	ug/l	90
84) 1,2,4-Trimethylbenzene	11.756	105	11824	2.723	ug/l	89
95) Naphthalene	13.774	128	374639	60.984	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071524\
Data File : VX042324.D
Acq On : 15 Jul 2024 14:01
Operator : JC/MD
Sample : P3224-01 20X
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
380AVEC-21

Quant Time: Jul 16 05:39:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

