

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071524\
 Data File : VX042325.D
 Acq On : 15 Jul 2024 14:25
 Operator : JC/MD
 Sample : P3224-25 20X
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 356STEVEN-13

Quant Time: Jul 16 05:39:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	154841	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	239186	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226097	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110345	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	90856	50.885	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.780%
35) Dibromofluoromethane	5.391	113	87427	51.514	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.020%
50) Toluene-d8	8.653	98	293782	49.274	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.540%
62) 4-Bromofluorobenzene	11.085	95	108011	49.145	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.280%
Target Compounds						
						Qvalue
95) Naphthalene	13.780	128	5354	1.414	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071524\
Data File : VX042325.D
Acq On : 15 Jul 2024 14:25
Operator : JC/MD
Sample : P3224-25 20X
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
356STEVEN-13

Quant Time: Jul 16 05:39:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

