

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071524\
 Data File : VX042329.D
 Acq On : 15 Jul 2024 16:37
 Operator : JC/MD
 Sample : P3224-09
 Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 400STEVEN-2

Quant Time: Jul 16 05:40:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	137850	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	216532	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	193552	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93206	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86058	54.139	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.280%	
35) Dibromofluoromethane	5.397	113	77731	50.593	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.180%	
50) Toluene-d8	8.647	98	260227	48.213	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 96.420%	
62) 4-Bromofluorobenzene	11.085	95	96290	48.396	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.800%	
Target Compounds						
					Qvalue	
5) Bromomethane	1.599	94	779	1.195	ug/l	89
68) m/p-Xylenes	10.305	106	8767	3.443	ug/l	98
69) o-Xylene	10.640	106	4497	1.784	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	6492	1.375	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	12096	2.554	ug/l	96
95) Naphthalene	13.774	128	319077	47.782	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071524\
Data File : VX042329.D
Acq On : 15 Jul 2024 16:37
Operator : JC/MD
Sample : P3224-09
Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
400STEVEN-2

Quant Time: Jul 16 05:40:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

