

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071524\
 Data File : VX042330.D
 Acq On : 15 Jul 2024 17:01
 Operator : JC/MD
 Sample : P3224-17 20X
 Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 356STEVEN-7

Quant Time: Jul 16 05:40:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	151540	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	234076	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219567	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108610	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84344	48.267	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.540%
35) Dibromofluoromethane	5.391	113	82030	49.389	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.780%
50) Toluene-d8	8.653	98	278321	47.700	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.400%
62) 4-Bromofluorobenzene	11.079	95	102274	47.550	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.100%
Target Compounds						
					Qvalue	
95) Naphthalene	13.780	128	4361	1.300	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071524\
Data File : VX042330.D
Acq On : 15 Jul 2024 17:01
Operator : JC/MD
Sample : P3224-17 20X
Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
356STEVEN-7

Quant Time: Jul 16 05:40:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

