

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071525\
 Data File : VX046990.D
 Acq On : 15 Jul 2025 11:00
 Operator : JC/MD
 Sample : VX0715WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0715WBL01

Quant Time: Jul 16 02:21:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	334020	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	564806	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	510782	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	265478	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	230225	52.173	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.340%
35) Dibromofluoromethane	5.397	113	188128	49.069	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.140%
50) Toluene-d8	8.653	98	652531	48.240	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.480%
62) 4-Bromofluorobenzene	11.079	95	269032	52.331	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071525\
Data File : VX046990.D
Acq On : 15 Jul 2025 11:00
Operator : JC/MD
Sample : VX0715WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0715WBL01

Quant Time: Jul 16 02:21:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

