

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071525\
 Data File : VX046998.D
 Acq On : 15 Jul 2025 13:54
 Operator : JC/MD
 Sample : Q2579-01RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-A2-13-GRE

Manual Integrations
 APPROVED

Reviewed By : John Carlone 07/16/2025
 Supervised By : Mahesh Dadoda 07/16/2025

Quant Time: Jul 16 02:26:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	352031	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	597406	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	550618	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	287789	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	250512	53.866	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 107.740%			
35) Dibromofluoromethane	5.409	113	447m	0.110	ug/l	0.01
Spiked Amount 50.000	Range 75 - 124		Recovery = 0.220%#			
50) Toluene-d8	8.653	98	696313	48.667	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 97.340%			
62) 4-Bromofluorobenzene	11.079	95	285624	52.526	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 105.060%			
Target Compounds						Qvalue
16) Acetone	2.392	43	100079	69.278	ug/l	98
20) Methylene Chloride	2.806	84	22311	5.223	ug/l	94
95) Naphthalene	13.774	128	1658311	101.431	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071525\
Data File : VX046998.D
Acq On : 15 Jul 2025 13:54
Operator : JC/MD
Sample : Q2579-01RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-A2-13-GRE

Manual Integrations
APPROVED

Reviewed By : John Carlone 07/16/2025
Supervised By : Mahesh Dadoda 07/16/2025

Quant Time: Jul 16 02:26:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

