

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX071619\
 Data File : VX010907.D
 Acq On : 16 Jul 2019 15:32
 Operator : JC/SP
 Sample : K3591-07 0.2PPB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2019

Quant Time: Jul 17 05:57:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 17 05:51:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	227267	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	339712	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	305065	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	146856	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	115121	47.61	ug/l	0.00
Spiked Amount			Recovery	=	95.22%	
35) Dibromofluoromethane	5.49	113	101681	47.16	ug/l	-0.01
Spiked Amount			Recovery	=	94.32%	
50) Toluene-d8	8.71	98	389190	48.12	ug/l	0.00
Spiked Amount			Recovery	=	96.24%	
62) 4-Bromofluorobenzene	11.13	95	134627	45.29	ug/l	0.00
Spiked Amount			Recovery	=	90.58%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.33	50	905	0.433	ug/l	98
4) Vinyl Chloride	1.41	62	637	0.299	ug/l #	76
5) Bromomethane	1.65	94	1311	1.091	ug/l	85
6) Chloroethane	1.73	64	790	0.583	ug/l	89
7) Trichlorofluoromethane	1.93	101	873	0.242	ug/l #	79
8) Diethyl Ether	2.19	74	517	0.416	ug/l	88
9) 1,1,2-Trichlorotrifluoroet	2.38	101	550	0.267	ug/l #	77
10) Methyl Iodide	2.51	142	897	0.332	ug/l #	95
11) Tert butyl alcohol	3.02	59	2535	4.502	ug/l #	89
12) 1,1-Dichloroethene	2.37	96	665	0.329	ug/l	90
13) Acrolein	2.29	56	886	2.723	ug/l #	70
14) Allyl chloride	2.73	41	1050	0.357	ug/l #	82
15) Acrylonitrile	3.14	53	1804	1.586	ug/l	86
16) Acetone	2.44	43	1667	1.527	ug/l	96
17) Carbon Disulfide	2.57	76	3488	0.672	ug/l #	92
18) Methyl Acetate	2.77	43	1155	0.465	ug/l #	86
19) Methyl tert-butyl Ether	3.19	73	1669	0.256	ug/l #	91
20) Methylene Chloride	2.85	84	1024	0.442	ug/l #	75
21) trans-1,2-Dichloroethene	3.17	96	989	0.460	ug/l #	70
22) Diisopropyl ether	3.85	45	1752	0.283	ug/l #	93
23) Vinyl Acetate	3.82	43	5378	1.024	ug/l #	87
24) 1,1-Dichloroethane	3.70	63	1036	0.288	ug/l #	82
25) 2-Butanone	4.68	43	1993	1.271	ug/l #	79
26) 2,2-Dichloropropane	4.57	77	1003	0.339	ug/l	94
27) cis-1,2-Dichloroethene	4.60	96	1038	0.418	ug/l	86
28) Bromochloromethane	5.02	49	500	0.295	ug/l #	78
29) Tetrahydrofuran	5.14	42	1163	1.171	ug/l #	61
30) Chloroform	5.20	83	833	0.214	ug/l	87
32) 1,1,1-Trichloroethane	5.48	97	859	0.255	ug/l #	35
36) 1,1-Dichloropropene	5.81	75	1045	0.374	ug/l #	91
38) Carbon Tetrachloride	5.78	117	862	0.296	ug/l #	88
39) Methylcyclohexane	7.46	83	1029	0.284	ug/l #	86
40) Benzene	6.13	78	2326	0.272	ug/l	94
41) Methacrylonitrile	5.05	41	493	0.304	ug/l #	57

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071619\
 Data File : VX010907.D
 Acq On : 16 Jul 2019 15:32
 Operator : JC/SP
 Sample : K3591-07 0.2PPB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2019

Quant Time: Jul 17 05:57:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 17 05:51:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 1,2-Dichloroethane	6.19	62	1191	0.401	ug/l	83
43) Isopropyl Acetate	6.45	43	1365	0.291	ug/l #	92
44) Trichloroethene	7.22	130	863	0.341	ug/l	91
45) 1,2-Dichloropropane	7.51	63	638	0.293	ug/l #	87
46) Dibromomethane	7.66	93	487	0.313	ug/l	91
47) Bromodichloromethane	7.89	83	737	0.266	ug/l #	87
48) Methyl methacrylate	7.77	41	865	0.378	ug/l #	62
49) 1,4-Dioxane	7.74	88	298	4.655	ug/l #	80
51) 4-Methyl-2-Pentanone	8.64	43	3345	1.098	ug/l	97
52) Toluene	8.78	92	1655	0.295	ug/l	100
53) t-1,3-Dichloropropene	9.04	75	718	0.241	ug/l	94
54) cis-1,3-Dichloropropene	8.43	75	725	0.219	ug/l #	86
55) 1,1,2-Trichloroethane	9.22	97	514	0.222	ug/l #	81
56) Ethyl methacrylate	9.18	69	677	0.200	ug/l #	30
57) 1,3-Dichloropropane	9.37	76	854	0.233	ug/l #	67
58) 2-Chloroethyl Vinyl ether	8.31	63	1424	0.882	ug/l #	87
59) 2-Hexanone	9.50	43	2243	0.950	ug/l	85
61) 1,2-Dibromoethane	9.66	107	695	0.283	ug/l	98
64) Tetrachloroethene	9.34	164	791	0.330	ug/l #	79
65) Chlorobenzene	10.13	112	1910	0.305	ug/l	97
66) 1,1,1,2-Tetrachloroethane	10.22	131	535	0.243	ug/l #	62
67) Ethyl Benzene	10.25	91	2714	0.254	ug/l	90
68) m/p-Xylenes	10.36	106	2031	0.496	ug/l	97
69) o-Xylene	10.70	106	1149	0.288	ug/l	75
70) Styrene	10.71	104	1549	0.233	ug/l	91
73) Isopropylbenzene	11.02	105	2628	0.265	ug/l	89
74) N-ethyl acetate	10.90	43	1038	0.274	ug/l	96
75) 1,1,2,2-Tetrachloroethane	11.27	83	842	0.259	ug/l #	83
76) 1,2,3-Trichloropropane	11.13	75	66677	25.160	ug/l #	35
77) Bromobenzene	11.26	156	968	0.354	ug/l	83
78) n-propylbenzene	11.36	91	2969	0.271	ug/l	97
79) 2-Chlorotoluene	11.42	91	1792	0.272	ug/l	92
80) 1,3,5-Trimethylbenzene	11.51	105	2118	0.255	ug/l	96
81) trans-1,4-Dichloro-2-buten	11.13	75	66677	69.228	ug/l #	11
82) 4-Chlorotoluene	11.51	91	2308	0.301	ug/l	91
83) tert-Butylbenzene	11.77	119	1902	0.234	ug/l	94
84) 1,2,4-Trimethylbenzene	11.80	105	2069	0.248	ug/l	99
85) sec-Butylbenzene	11.94	105	2262	0.237	ug/l	97
86) p-Isopropyltoluene	12.07	119	2170	0.246	ug/l #	71
87) 1,3-Dichlorobenzene	12.02	146	1586	0.334	ug/l	93
88) 1,4-Dichlorobenzene	12.02	146	1586	0.326	ug/l	91
89) n-Butylbenzene	12.39	91	1940	0.257	ug/l	97
90) Hexachloroethane	12.60	117	302	0.217	ug/l	88
91) 1,2-Dichlorobenzene	12.39	146	1373	0.292	ug/l	87
92) 1,2-Dibromo-3-Chloropropan	13.00	75	163	0.240	ug/l	73
93) 1,2,4-Trichlorobenzene	13.64	180	986	0.309	ug/l	80
94) Hexachlorobutadiene	13.78	225	495	0.313	ug/l	88
95) Naphthalene	13.83	128	2665	0.283	ug/l	96
96) 1,2,3-Trichlorobenzene	14.02	180	1150	0.362	ug/l	96

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071619\
Data File : VX010907.D
Acq On : 16 Jul 2019 15:32
Operator : JC/SP
Sample : K3591-07 0.2PPB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LOD-MDL-WATER-01-QT3-2019

Quant Time: Jul 17 05:57:08 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 17 05:51:36 2019
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
--------------------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX071619\
 Data File : VX010907.D
 Acq On : 16 Jul 2019 15:32
 Operator : JC/SP
 Sample : K3591-07 0.2PPB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2019

Quant Time: Jul 17 05:57:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 17 05:51:36 2019
 Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 738

Delta R.T. -0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :


LOD-MDL-WATER-01-QT3-2019

Tot Ion:168 Resp: 227267

Ion Ratio Lower Upper




168 100

99 50.9 37.6 56.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#3

Chloromethane

Concen: 0.433 ug/l

RT: 1.33 min Scan# 29

Delta R.T. -0.00 min

Lab File: VX010907.D


Acq: 16 Jul 2019 15:32

Tgt Ion: 50 Resp: 905

Ion Ratio Lower Upper



50 100


52 34.1 26.4 39.6



Abundance Ion 49.90 (49.60 to 50.60): VX0

Ion 51.90 (51.60 to 52.60): VX0





#4

Vinyl Chloride

Concen: 0.299 ug/l

RT: 1.41 min Scan# 42

Delta R.T. -0.00 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :


LOD-MDL-WATER-01-QT3-2019

Tot Ion: 62 Resp: 637

Ion Ratio Lower Upper

62 100

64 18.8 25.7 38.5#



Abundance Ion 61.90 (61.60 to 62.60): VX0

Ion 63.90 (63.60 to 64.60): VX0

600

500

400

300

200

100



0

1.35

1.40

1.45

Time-->



#5

Bromomethane

Concen: 1.091 ug/l

RT: 1.65 min Scan# 81

Delta R.T. 0.01 min

Lab File: VX010907.D


Acq: 16 Jul 2019 15:32

Tgt Ion: 94 Resp: 1311

Ion Ratio Lower Upper

94 100

96 80.1 75.6 113.4



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0

800

600

400

200

0


1.60


1.65

1.70

1.75

Time-->





#6

Chloroethane

Concen: 0.583 ug/l

RT: 1.73 min Scan# 95

Delta R.T. 0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :


LOD-MDL-WATER-01-QT3-2019

Tot Ion: 64 Resp: 790




Ion Ratio Lower Upper

64 100

66 27.7 27.0 40.6



Abundance Ion 63.90 (63.60 to 64.60): VX0



#7

Trichlorofluoromethane

Concen: 0.242 ug/l

RT: 1.93 min Scan# 128

Delta R.T. 0.01 min

Lab File: VX010907.D


Acq: 16 Jul 2019 15:32

Tgt Ion: 101 Resp: 873



Ion Ratio Lower Upper


101 100

103 48.9 52.6 78.8#



Abundance Ion 100.90 (100.60 to 101.60): V





#8

Diethyl Ether

Concen: 0.416 ug/l

RT: 2.19 min Scan# 170

Delta R.T. -0.00 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :





LOD-MDL-WATER-01-QT3-2019

Tot Ion: 74 Resp: 517

Ion Ratio Lower Upper

74 100

45 79.9 45.5 136.5



#9

1,1,2-Trichlorotrifluoroethane

Concen: 0.267 ug/l

RT: 2.38 min Scan# 201

Delta R.T. -0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32




Tgt Ion: 101 Resp: 550


Ion Ratio Lower Upper

101 100

85 54.7 34.6 52.0#

151 110.9 69.6 104.4#









#10
Methvl Iodide
Concen: 0.332 ug/l
RT: 2.51 min Scan# 223
Delta R.T. 0.01 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32

Instrument :
MSVOA_X
ClientSampleId :
LOD-MDL-WATER-01-QT3-2019

Tot Ion:142 Resp: 897
Ion Ratio Lower Upper
142 100
127 45.0 34.2 51.4
141 18.2 11.7 17.5#




Abundance Ion 141.80 (141.50 to 142.50): V
Ion 126.80 (126.50 to 127.50): V
Ion 140.80 (140.50 to 141.50): V






#11
Tert butyl alcohol
Concen: 4.502 ug/l
RT: 3.02 min Scan# 306
Delta R.T. -0.03 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32

Tgt Ion: 59 Resp: 2535
Ion Ratio Lower Upper
59 100
57 5.8 7.9 11.9#



Abundance Ion 59.00 (58.70 to 59.70): VX0
Ion 57.00 (56.70 to 57.70): VX0





#12

1,1-Dichloroethene

Concen: 0.329 ug/l

RT: 2.37 min Scan# 200

Delta R.T. -0.00 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :

LOD-MDL-WATER-01-QT3-2019


Tot Ion: 96 Resp: 665

Ion Ratio Lower Upper

96 100

61 158.3 115.0 172.4




98 70.4 52.2 78.2



Abundance Ion 95.90 (95.60 to 96.60): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 97.90 (97.60 to 98.60): VX0



#13

Acrolein

Concen: 2.723 ug/l

RT: 2.29 min Scan# 187

Delta R.T. -0.00 min

Lab File: VX010907.D


Acq: 16 Jul 2019 15:32

Tgt Ion: 56 Resp: 886

Ion Ratio Lower Upper



56 100


55 94.7 55.9 83.9#



Abundance Ion 56.00 (55.70 to 56.70): VX0

Ion 55.00 (54.70 to 55.70): VX0





#14

Allvl chloride

Concen: 0.357 ug/l

RT: 2.73 min Scan# 258

Delta R.T. -0.00 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :

LOD-MDL-WATER-01-QT3-2019


Tot Ion: 41 Resp: 1050

Ion Ratio Lower Upper




41 100

39 51.9 53.6 80.4#

76 26.9 29.7 44.5#



Abundance Ion 41.00 (40.70 to 41.70): VX0
Ion 39.00 (38.70 to 39.70): VX0
Ion 75.90 (75.60 to 76.60): VX0



#15

Acrylonitrile

Concen: 1.586 ug/l

RT: 3.14 min Scan# 325

Delta R.T. -0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32


Tgt Ion: 53 Resp: 1804

Ion Ratio Lower Upper



53 100


52 96.8 66.5 99.7

51 42.3 28.2 42.4



Abundance Ion 53.00 (52.70 to 53.70): VX0
Ion 52.00 (51.70 to 52.70): VX0
Ion 51.00 (50.70 to 51.70): VX0





#16

Acetone

Concen: 1.527 ug/l

RT: 2.44 min Scan# 211

Delta R.T. -0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :





LOD-MDL-WATER-01-QT3-2019

Tot Ion: 43 Resp: 1667

Ion Ratio Lower Upper

43 100

58 32.3 27.7 41.5



#17

Carbon Disulfide

Concen: 0.672 ug/l

RT: 2.57 min Scan# 232

Delta R.T. -0.00 min

Lab File: VX010907.D




Acq: 16 Jul 2019 15:32


Tgt Ion: 76 Resp: 3488

Ion Ratio Lower Upper

76 100

78 11.9 7.4 11.0#









#18
Methyl Acetate
Concen: 0.465 ug/l
RT: 2.77 min Scan# 265
Delta R.T. -0.01 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32




Instrument :
MSVOA_X
ClientSampleId :
LOD-MDL-WATER-01-QT3-2019


Tot Ion: 43 Resp: 1155
Ion Ratio Lower Upper
43 100
74 19.9 21.5 32.3#



#19
Methyl tert-butyl Ether
Concen: 0.256 ug/l
RT: 3.19 min Scan# 334
Delta R.T. -0.01 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32

Tgt Ion: 73 Resp: 1669
Ion Ratio Lower Upper
73 100
57 26.0 17.3 25.9#









#20
Methvlene Chloride
Concen: 0.442 ug/l
RT: 2.85 min Scan# 279
Delta R.T. -0.00 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32




Instrument :
MSVOA_X
ClientSampleId :
LOD-MDL-WATER-01-QT3-2019


Tot Ion: 84 Resp: 1024
Ion Ratio Lower Upper
84 100
49 156.6 89.9 134.9#
51 41.6 27.8 41.6
86 67.0 51.8 77.6



#21
trans-1,2-Dichloroethene
Concen: 0.460 ug/l
RT: 3.17 min Scan# 331
Delta R.T. 0.01 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32

Tgt Ion: 96 Resp: 989
Ion Ratio Lower Upper
96 100
61 77.5 104.2 156.2#
98 66.2 52.2 78.4





#22

Diisopropyl ether

Concen: 0.283 ug/l

RT: 3.85 min Scan# 443

Delta R.T. -0.00 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :

LOD-MDL-WATER-01-QT3-2019

Tot Ion: 45 Resp: 1752





Ion Ratio Lower Upper

45 100

43 83.4 62.3 93.5

87 33.2 25.0 37.4

59 20.8 9.9 14.9#



#23

Vinyl Acetate

Concen: 1.024 ug/l

RT: 3.82 min Scan# 437

Delta R.T. -0.00 min

Lab File: VX010907.D




Acq: 16 Jul 2019 15:32


Tgt Ion: 43 Resp: 5378

Ion Ratio Lower Upper

43 100

86 6.8 9.4 14.0#





#24

1,1-Dichloroethane

Concen: 0.288 ug/l

RT: 3.70 min Scan# 418

Delta R.T. 0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :

LOD-MDL-WATER-01-QT3-2019


Tot Ion: 63 Resp: 1036

Ion Ratio Lower Upper

63 100

98 0.0 3.6 10.9#


100 0.0 2.4 7.2#





Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 97.90 (97.60 to 98.60): VX0

Ion 99.90 (99.60 to 100.60): VX0



Time-->



#25

2-Butanone

Concen: 1.271 ug/l

RT: 4.68 min Scan# 578

Delta R.T. 0.01 min

Lab File: VX010907.D


Acq: 16 Jul 2019 15:32

Tgt Ion: 43 Resp: 1993

Ion Ratio Lower Upper


43 100

72 39.8 23.0 34.6#





Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 72.00 (71.70 to 72.70): VX0



Time-->





#26

2,2-Dichloropropane

Concen: 0.339 ug/l

RT: 4.57 min Scan# 561

Delta R.T. -0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :


LOD-MDL-WATER-01-QT3-2019

Tot Ion: 77 Resp: 1003

Ion Ratio Lower Upper

77 100

97 21.9 12.6 37.6



Abundance Ion 76.90 (76.60 to 77.60): VX0

Ion 96.90 (96.60 to 97.60): VX0

4.57



300

200

100

0

Time-->



#27

cis-1,2-Dichloroethene

Concen: 0.418 ug/l

RT: 4.60 min Scan# 565

Delta R.T. -0.00 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32


Tgt Ion: 96 Resp: 1038

Ion Ratio Lower Upper

96 100

61 118.6 0.0 269.0

98 52.8 0.0 131.4



Abundance Ion 95.90 (95.60 to 96.60): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 97.90 (97.60 to 98.60): VX0

600

500

400


300


200

100

0

Time-->





#28

Bromochloromethane

Concen: 0.295 ug/l

RT: 5.02 min Scan# 634

Delta R.T. 0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :

LOD-MDL-WATER-01-QT3-2019


Tot Ion: 49 Resp: 500

Ion Ratio Lower Upper

49 100

129 0.0 0.0 5.2




130 74.4 76.9 115.3#



Abundance Ion 48.90 (48.60 to 49.60): VX0

Ion 128.80 (128.50 to 129.50): V

Ion 129.80 (129.50 to 130.50): V



#29

Tetrahydrofuran

Concen: 1.171 ug/l

RT: 5.14 min Scan# 654

Delta R.T. 0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32


Tgt Ion: 42 Resp: 1163

Ion Ratio Lower Upper

42 100

72 8.9 39.8 59.8#



71 57.5 37.0 55.4#




Abundance Ion 42.00 (41.70 to 42.70): VX0

Ion 72.00 (71.70 to 72.70): VX0

Ion 71.00 (70.70 to 71.70): VX0









#30
Chloroform
Concen: 0.214 ug/l
RT: 5.20 min Scan# 664
Delta R.T. -0.01 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32




Instrument :
MSVOA_X
ClientSampleId :
LOD-MDL-WATER-01-QT3-2019


Tot Ion: 83 Resp: 833
Ion Ratio Lower Upper
83 100
85 54.4 51.4 77.0



#32
1,1,1-Trichloroethane
Concen: 0.255 ug/l
RT: 5.48 min Scan# 710
Delta R.T. -0.01 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32

Tgt Ion: 97 Resp: 859
Ion Ratio Lower Upper
97 100
99 0.0 51.4 77.0#
61 65.1 33.0 49.6#





#33

1,2-Dichloroethane-d4

Concen: 47.614 ug/l

RT: 6.06 min Scan# 804

Delta R.T. -0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :





LOD-MDL-WATER-01-QT3-2019

Tot Ion: 65 Resp: 115121

Ion Ratio Lower Upper

65 100

67 53.7 0.0 107.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 935

Delta R.T. -0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32




Tgt Ion: 114 Resp: 339712


Ion Ratio Lower Upper

114 100

63 18.2 0.0 35.6

88 14.8 0.0 29.8









#35
Dibromofluoromethane
Concen: 47.157 ug/l
RT: 5.49 min Scan# 711
Delta R.T. -0.01 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32




Instrument :
MSVOA_X
ClientSampleId :
LOD-MDL-WATER-01-QT3-2019


Tot Ion:113 Resp: 101681
Ion Ratio Lower Upper
113 100
111 101.8 82.3 123.5
192 22.0 17.4 26.2



#36
1,1-Dichloropropene
Concen: 0.374 ug/l
RT: 5.81 min Scan# 763
Delta R.T. 0.01 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32

Tgt Ion: 75 Resp: 1045
Ion Ratio Lower Upper
75 100
110 39.1 19.9 59.7
77 21.2 25.0 37.4#





#38

Carbon Tetrachloride

Concen: 0.296 ug/l

RT: 5.78 min Scan# 759

Delta R.T. -0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :

LOD-MDL-WATER-01-QT3-2019


Tot Ion:117 Resp: 862

Ion Ratio Lower Upper

117 100

119 81.5 74.2 111.2




121 24.1 24.2 36.2#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#39

Methylcyclohexane

Concen: 0.284 ug/l

RT: 7.46 min Scan# 1034

Delta R.T. -0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32


Tgt Ion: 83 Resp: 1029

Ion Ratio Lower Upper

83 100

55 61.0 56.2 84.4



98 35.9 39.0 58.4#




Abundance Ion 83.00 (82.70 to 83.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 98.00 (97.70 to 98.70): VX0









#40
Benzene
Concen: 0.272 ug/l
RT: 6.13 min Scan# 816
Delta R.T. -0.01 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32




Instrument :
MSVOA_X
ClientSampleId :
LOD-MDL-WATER-01-QT3-2019


Tot Ion: 78 Resp: 2326
Ion Ratio Lower Upper
78 100
77 26.9 19.0 28.6



#41
Methacrylonitrile
Concen: 0.304 ug/l
RT: 5.05 min Scan# 639
Delta R.T. 0.01 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32

Tgt Ion: 41 Resp: 493
Ion Ratio Lower Upper
41 100
39 0.0 45.4 68.2#
67 54.2 65.9 98.9#
52 25.4 26.2 39.2#









#42
 1,2-Dichloroethane
 Concen: 0.401 ug/l
 RT: 6.19 min Scan# 826
 Delta R.T. -0.00 min
 Lab File: VX010907.D
 Acq: 16 Jul 2019 15:32




Instrument :
 MSVOA_X
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2019


Tot Ion: 62 Resp: 1191
 Ion Ratio Lower Upper
 62 100
 98 16.4 0.0 20.0



#43
 Isopropyl Acetate
 Concen: 0.291 ug/l
 RT: 6.45 min Scan# 868
 Delta R.T. -0.00 min
 Lab File: VX010907.D
 Acq: 16 Jul 2019 15:32

Tgt Ion: 43 Resp: 1365
 Ion Ratio Lower Upper
 43 100
 61 21.1 17.1 25.7
 87 7.2 12.0 18.0#





#44

Trichloroethene

Concen: 0.341 ug/l

RT: 7.22 min Scan# 995

Delta R.T. 0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :


LOD-MDL-WATER-01-QT3-2019

Tot Ion:130 Resp: 863




Ion Ratio Lower Upper

130 100

95 78.8 0.0 174.8



Abundance Ion 129.80 (129.50 to 130.50): V



#45

1,2-Dichloropropane

Concen: 0.293 ug/l

RT: 7.51 min Scan# 1043

Delta R.T. -0.00 min

Lab File: VX010907.D


Acq: 16 Jul 2019 15:32

Tgt Ion: 63 Resp: 638



Ion Ratio Lower Upper


63 100

65 37.5 24.4 36.6#



Abundance Ion 62.90 (62.60 to 63.60): VX0





#46

Dibromomethane

Concen: 0.313 ug/l

RT: 7.66 min Scan# 1067

Delta R.T. -0.00 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :

LOD-MDL-WATER-01-QT3-2019


Tot Ion: 93 Resp: 487

Ion Ratio Lower Upper

93 100

95 74.5 67.2 100.8




174 104.5 90.6 135.8



Abundance Ion 92.80 (92.50 to 93.50): VX0

Ion 94.80 (94.50 to 95.50): VX0

Ion 173.70 (173.40 to 174.40): V



#47

Bromodichloromethane

Concen: 0.266 ug/l

RT: 7.89 min Scan# 1105

Delta R.T. -0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32


Tgt Ion: 83 Resp: 737

Ion Ratio Lower Upper

83 100

85 56.6 52.5 78.7



127 0.0 7.8 11.8#




Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 84.90 (84.60 to 85.60): VX0

Ion 126.80 (126.50 to 127.50): V





#48

Methyl methacrylate

Concen: 0.378 ug/l

RT: 7.77 min Scan# 1086

Delta R.T. 0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :

LOD-MDL-WATER-01-QT3-2019


Tot Ion: 41 Resp: 865

Ion Ratio Lower Upper

41 100

69 36.2 74.4 111.6#




39 56.4 48.9 73.3



Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 69.00 (68.70 to 69.70): VX0

Ion 39.00 (38.70 to 39.70): VX0



#49

1,4-Dioxane

Concen: 4.655 ug/l

RT: 7.74 min Scan# 1081

Delta R.T. 0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32


Tgt Ion: 88 Resp: 298

Ion Ratio Lower Upper

88 100

43 55.4 22.9 34.3#



58 58.1 49.8 74.8




Abundance Ion 88.00 (87.70 to 88.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0









#50
Toluene-d8
Concen: 48.124 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. -0.00 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32




Instrument :
MSVOA_X
ClientSampleId :
LOD-MDL-WATER-01-QT3-2019


Tot Ion: 98 Resp: 389190
Ion Ratio Lower Upper
98 100
100 64.4 50.6 75.8



#51
4-Methyl-2-Pentanone
Concen: 1.098 ug/l
RT: 8.64 min Scan# 1228
Delta R.T. -0.00 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32

Tgt Ion: 43 Resp: 3345
Ion Ratio Lower Upper
43 100
58 42.1 32.2 48.4





#52

Toluene

Concen: 0.295 ug/l

RT: 8.78 min Scan# 1251

Delta R.T. -0.00 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :


LOD-MDL-WATER-01-QT3-2019

Tot Ion: 92 Resp: 1655

Ion Ratio Lower Upper




92 100

91 174.5 139.4 209.0



Abundance Ion 92.00 (91.70 to 92.70): VX0

Ion 91.00 (90.70 to 91.70): VX0



#53

t-1,3-Dichloropropene

Concen: 0.241 ug/l

RT: 9.04 min Scan# 1294

Delta R.T. 0.01 min

Lab File: VX010907.D


Acq: 16 Jul 2019 15:32

Tgt Ion: 75 Resp: 718

Ion Ratio Lower Upper



75 100


77 36.3 26.2 39.2



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 76.90 (76.60 to 77.60): VX0









#54
 cis-1,3-Dichloropropene
 Concen: 0.219 ug/l
 RT: 8.43 min Scan# 1194
 Delta R.T. -0.00 min
 Lab File: VX010907.D
 Acq: 16 Jul 2019 15:32




Instrument :
 MSVOA_X
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2019


Tot Ion: 75 Resp: 725
 Ion Ratio Lower Upper
 75 100
 77 48.4 26.1 39.1#
 39 48.4 36.6 54.8



#55
 1,1,2-Trichloroethane
 Concen: 0.222 ug/l
 RT: 9.22 min Scan# 1323
 Delta R.T. 0.01 min
 Lab File: VX010907.D
 Acq: 16 Jul 2019 15:32

Tgt Ion: 97 Resp: 514
 Ion Ratio Lower Upper
 97 100
 83 67.0 69.2 103.8#
 85 31.3 44.6 66.8#
 99 67.0 51.2 76.8





#56

Ethyl methacrylate

Concen: 0.200 ug/l

RT: 9.18 min Scan# 1316

Delta R.T. -0.00 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :

LOD-MDL-WATER-01-QT3-2019





Tot Ion: 69 Resp: 677

Ion Ratio Lower Upper

69 100

41 108.6 49.7 74.5#

39 88.0 29.6 44.4#



#57

1,3-Dichloropropane

Concen: 0.233 ug/l

RT: 9.37 min Scan# 1348

Delta R.T. 0.01 min

Lab File: VX010907.D




Acq: 16 Jul 2019 15:32


Tgt Ion: 76 Resp: 854

Ion Ratio Lower Upper

76 100

78 50.6 25.6 38.4#









#58
2-Chloroethyl Vinyl ether
Concen: 0.882 ug/l
RT: 8.31 min Scan# 1174
Delta R.T. -0.00 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32




Instrument :
MSVOA_X
ClientSampleId :
LOD-MDL-WATER-01-QT3-2019


Tot Ion: 63 Resp: 1424
Ion Ratio Lower Upper
63 100
106 36.9 24.0 36.0#



#59
2-Hexanone
Concen: 0.950 ug/l
RT: 9.50 min Scan# 1369
Delta R.T. 0.01 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32

Tgt Ion: 43 Resp: 2243
Ion Ratio Lower Upper
43 100
58 67.0 28.0 83.9





#61

1,2-Dibromoethane

Concen: 0.283 ug/l

RT: 9.66 min Scan# 1396

Delta R.T. -0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :


LOD-MDL-WATER-01-QT3-2019

Tot Ion:107 Resp: 695

Ion Ratio Lower Upper

107 100

109 95.7 75.4 113.0



Abundance Ion 106.90 (106.60 to 107.60): V

Ion 108.80 (108.50 to 109.50): V

9.66

400

300

200



100

0

9.65

9.70

Time-->



#62

4-Bromofluorobenzene

Concen: 45.289 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.00 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32


Tgt Ion: 95 Resp: 134627

Ion Ratio Lower Upper

95 100

174 92.4 0.0 180.6

176 89.1 0.0 173.8



Abundance Ion 94.90 (94.60 to 95.60): VX0

150000 Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

11.13

100000

50000

0


11.05


11.10

11.15

11.20

Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. -0.00 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32




Instrument :
MSVOA_X
ClientSampleId :
LOD-MDL-WATER-01-QT3-2019


Tot Ion:117 Resp: 305065
Ion Ratio Lower Upper
117 100
82 51.3 42.6 63.8
119 31.3 25.9 38.9



#64
Tetrachloroethene
Concen: 0.330 ug/l
RT: 9.34 min Scan# 1342
Delta R.T. -0.00 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32

Tgt Ion:164 Resp: 791
Ion Ratio Lower Upper
164 100
166 93.4 100.8 151.2#
129 70.2 72.6 109.0#
131 76.7 67.7 101.5









#65
Chlorobenzene
Concen: 0.305 ug/l
RT: 10.13 min Scan# 1473
Delta R.T. -0.00 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32




Instrument :
MSVOA_X
ClientSampleId :
LOD-MDL-WATER-01-QT3-2019


Tot Ion:112 Resp: 1910
Ion Ratio Lower Upper
112 100
114 31.4 26.2 39.4



#66
1,1,1,2-Tetrachloroethane
Concen: 0.243 ug/l
RT: 10.22 min Scan# 1487
Delta R.T. -0.00 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32

Tgt Ion:131 Resp: 535
Ion Ratio Lower Upper
131 100
133 86.4 47.1 141.4
119 0.0 31.9 95.7#









#67
Ethyl Benzene
Concen: 0.254 ug/l
RT: 10.25 min Scan# 1492
Delta R.T. -0.00 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32




Instrument :
MSVOA_X
ClientSampleId :
LOD-MDL-WATER-01-QT3-2019


Tot Ion: 91 Resp: 2714
Ion Ratio Lower Upper
91 100
106 26.1 25.4 38.2



#68
m/p-Xylenes
Concen: 0.496 ug/l
RT: 10.36 min Scan# 1510
Delta R.T. -0.00 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32

Tgt Ion: 106 Resp: 2031
Ion Ratio Lower Upper
106 100
91 195.9 160.1 240.1









#69
o-Xylene
Concen: 0.288 ug/l
RT: 10.70 min Scan# 1566
Delta R.T. 0.01 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32

Instrument :
MSVOA_X
ClientSampleId :
LOD-MDL-WATER-01-QT3-2019

Tot Ion:106 Resp: 1149
Ion Ratio Lower Upper
106 100
91 170.5 105.1 315.3




Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VX0






#70
Styrene
Concen: 0.233 ug/l
RT: 10.71 min Scan# 1568
Delta R.T. 0.01 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32

Tgt Ion:104 Resp: 1549
Ion Ratio Lower Upper
104 100
78 54.7 38.8 58.2
103 61.9 44.2 66.4



Abundance Ion 104.00 (103.70 to 104.70): V
Ion 78.00 (77.70 to 78.70): VX0
Ion 103.00 (102.70 to 103.70): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. -0.00 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :

LOD-MDL-WATER-01-QT3-2019


Tot Ion:152 Resp: 146856

Ion Ratio Lower Upper

152 100

115 56.2 39.7 119.1




150 157.2 0.0 345.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#73

Isopropylbenzene

Concen: 0.265 ug/l

RT: 11.02 min Scan# 1618

Delta R.T. -0.00 min

Lab File: VX010907.D


Acq: 16 Jul 2019 15:32

Tgt Ion:105 Resp: 2628

Ion Ratio Lower Upper



105 100


120 33.4 13.7 41.1



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#74

N-amvl acetate

Concen: 0.274 ug/l

RT: 10.90 min Scan# 1598

Delta R.T. -0.00 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :

LOD-MDL-WATER-01-QT3-2019

Tot Ion: 43 Resp: 1038



Ion Ratio Lower Upper

43 100

70 45.0 34.7 52.1

55 22.7 19.1 28.7

61 28.7 19.9 29.9




Abundance Ion 43.00 (42.70 to 43.70): VX0


Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0



Time-->



#75

1,1,2,2-Tetrachloroethane

Concen: 0.259 ug/l

RT: 11.27 min Scan# 1659

Delta R.T. -0.00 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32



Tgt Ion: 83 Resp: 842

Ion Ratio Lower Upper

83 100

131 0.0 5.5 16.4#


85 76.8 32.3 96.9




Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 130.80 (130.50 to 131.50): V

Ion 84.90 (84.60 to 85.60): VX0







Time-->



#76
1,2,3-Trichloropropane
Concen: 25.160 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. -0.16 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32




Instrument :
MSVOA_X
ClientSampleId :
LOD-MDL-WATER-01-QT3-2019


Tot Ion: 75 Resp: 66677
Ion Ratio Lower Upper
75 100
77 1.3 21.3 64.0#



#77
Bromobenzene
Concen: 0.354 ug/l
RT: 11.26 min Scan# 1657
Delta R.T. -0.00 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32

Tgt Ion: 156 Resp: 968
Ion Ratio Lower Upper
156 100
77 120.5 68.7 206.1
158 74.2 47.9 143.7





#78

n-propylbenzene

Concen: 0.271 ug/l

RT: 11.36 min Scan# 1674

Delta R.T. -0.00 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :


LOD-MDL-WATER-01-QT3-2019

Tot Ion: 91 Resp: 2969

Ion Ratio Lower Upper


91 100

120 26.2 12.3 36.8





Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 120.00 (119.70 to 120.70): V



Time-->



#79

2-Chlorotoluene

Concen: 0.272 ug/l

RT: 11.42 min Scan# 1684

Delta R.T. -0.00 min

Lab File: VX010907.D


Acq: 16 Jul 2019 15:32

Tgt Ion: 91 Resp: 1792

Ion Ratio Lower Upper


91 100

126 40.7 18.0 54.0





Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 125.90 (125.60 to 126.60): V



Time-->









#80
 1,3,5-Trimethylbenzene
 Concen: 0.255 ug/l
 RT: 11.51 min Scan# 1698
 Delta R.T. -0.00 min
 Lab File: VX010907.D
 Acq: 16 Jul 2019 15:32




Instrument :
 MSVOA_X
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2019


Tot Ion:105 Resp: 2118
 Ion Ratio Lower Upper
 105 100
 120 51.9 24.6 73.8



#81
 trans-1,4-Dichloro-2-butene
 Concen: 69.228 ug/l
 RT: 11.13 min Scan# 1637
 Delta R.T. 0.06 min
 Lab File: VX010907.D
 Acq: 16 Jul 2019 15:32

Tgt Ion: 75 Resp: 66677
 Ion Ratio Lower Upper
 75 100
 53 0.1 76.7 115.1#
 89 0.0 38.5 57.7#









#82
4-Chlorotoluene
Concen: 0.301 ug/l
RT: 11.51 min Scan# 1699
Delta R.T. 0.01 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32




Instrument :
MSVOA_X
ClientSampleId :
LOD-MDL-WATER-01-QT3-2019


Tot Ion: 91 Resp: 2308
Ion Ratio Lower Upper
91 100
126 36.9 15.9 47.7



#83
tert-Butylbenzene
Concen: 0.234 ug/l
RT: 11.77 min Scan# 1741
Delta R.T. -0.00 min
Lab File: VX010907.D
Acq: 16 Jul 2019 15:32

Tgt Ion: 119 Resp: 1902
Ion Ratio Lower Upper
119 100
91 53.8 28.6 85.8
134 28.3 11.6 34.8





#84

1,2,4-Trimethylbenzene

Concen: 0.248 ug/l

RT: 11.80 min Scan# 1747

Delta R.T. -0.00 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :


LOD-MDL-WATER-01-QT3-2019

Tot Ion:105 Resp: 2069

Ion Ratio Lower Upper




105 100

120 45.0 22.9 68.5



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#85

sec-Butylbenzene

Concen: 0.237 ug/l

RT: 11.94 min Scan# 1770

Delta R.T. -0.00 min

Lab File: VX010907.D


Acq: 16 Jul 2019 15:32

Tgt Ion:105 Resp: 2262

Ion Ratio Lower Upper



105 100


134 22.3 10.4 31.4



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V









#86
 p-Isopropyltoluene
 Concen: 0.246 ug/l
 RT: 12.07 min Scan# 1790
 Delta R.T. 0.01 min
 Lab File: VX010907.D
 Acq: 16 Jul 2019 15:32

Instrument :
 MSVOA_X
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2019

Tot Ion:119 Resp: 2170
 Ion Ratio Lower Upper
 119 100
 134 35.1 13.4 40.1
 91 45.2 11.8 35.4#




Abundance Ion 119.00 (118.70 to 119.70): V
 Ion 134.00 (133.70 to 134.70): V
 Ion 91.00 (90.70 to 91.70): VX0






#87
 1,3-Dichlorobenzene
 Concen: 0.334 ug/l
 RT: 12.02 min Scan# 1783
 Delta R.T. -0.00 min
 Lab File: VX010907.D
 Acq: 16 Jul 2019 15:32

Tgt Ion:146 Resp: 1586
 Ion Ratio Lower Upper
 146 100
 111 43.4 18.9 56.9
 148 69.1 32.2 96.6



Abundance Ion 145.90 (145.60 to 146.60): V
 Ion 110.90 (110.60 to 111.60): V
 Ion 147.90 (147.60 to 148.60): V





#88

1,4-Dichlorobenzene

Concen: 0.326 ug/l

RT: 12.02 min Scan# 1783

Delta R.T. -0.07 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :

LOD-MDL-WATER-01-QT3-2019


Tot Ion:146 Resp: 1586

Ion Ratio Lower Upper

146 100

111 43.4 18.6 55.8




148 69.1 31.6 94.7



Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V



#89

n-Butylbenzene

Concen: 0.257 ug/l

RT: 12.39 min Scan# 1843

Delta R.T. 0.01 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32


Tgt Ion: 91 Resp: 1940

Ion Ratio Lower Upper

91 100

92 55.2 26.2 78.5



134 28.7 13.8 41.3




Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 92.00 (91.70 to 92.70): VX0

Ion 134.00 (133.70 to 134.70): V









#90
 Hexachloroethane
 Concen: 0.217 ug/l
 RT: 12.60 min Scan# 1877
 Delta R.T. 0.01 min
 Lab File: VX010907.D
 Acq: 16 Jul 2019 15:32




Instrument :
 MSVOA_X
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2019


Tot Ion:117 Resp: 302
 Ion Ratio Lower Upper
 117 100
 201 75.8 43.3 129.8



#91
 1,2-Dichlorobenzene
 Concen: 0.292 ug/l
 RT: 12.39 min Scan# 1843
 Delta R.T. -0.00 min
 Lab File: VX010907.D
 Acq: 16 Jul 2019 15:32

Tgt Ion:146 Resp: 1373
 Ion Ratio Lower Upper
 146 100
 111 47.9 19.9 59.7
 148 73.8 32.0 96.2









#92
 1,2-Dibromo-3-Chloropropane
 Concen: 0.240 ug/l
 RT: 13.00 min Scan# 1943
 Delta R.T. -0.00 min
 Lab File: VX010907.D
 Acq: 16 Jul 2019 15:32




Instrument :
 MSVOA_X
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2019


Tot Ion: 75 Resp: 163
 Ion Ratio Lower Upper
 75 100
 155 69.3 48.7 146.1
 157 95.7 62.5 187.6



#93
 1,2,4-Trichlorobenzene
 Concen: 0.309 ug/l
 RT: 13.64 min Scan# 2048
 Delta R.T. -0.01 min
 Lab File: VX010907.D
 Acq: 16 Jul 2019 15:32

Tgt Ion: 180 Resp: 986
 Ion Ratio Lower Upper
 180 100
 182 119.6 47.1 141.3
 145 29.6 14.8 44.4





#94

Hexachlorobutadiene

Concen: 0.313 ug/l

RT: 13.78 min Scan# 2071

Delta R.T. -0.00 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32

Instrument :

MSVOA_X

ClientSampleId :

LOD-MDL-WATER-01-QT3-2019


Tot Ion:225 Resp: 495

Ion Ratio Lower Upper

225 100

223 76.6 31.9 95.9




227 58.4 32.5 97.4



Abundance Ion 224.70 (224.40 to 225.40): V

Ion 222.70 (222.40 to 223.40): V

Ion 226.70 (226.40 to 227.40): V



#95

Naphthalene

Concen: 0.283 ug/l

RT: 13.83 min Scan# 2080

Delta R.T. -0.00 min

Lab File: VX010907.D

Acq: 16 Jul 2019 15:32


Tgt Ion:128 Resp: 2665

Ion Ratio Lower Upper

128 100

127 14.9 10.4 15.6



129 9.9 8.9 13.3




Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V






#96
 1,2,3-Trichlorobenzene
 Concen: 0.362 ug/l
 RT: 14.02 min Scan# 2111
 Delta R.T. 0.01 min
 Lab File: VX010907.D
 Acq: 16 Jul 2019 15:32

Instrument :
 MSVOA_X
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2019

Tot Ion	Ratio	Lower	Upper
180	100		
182	90.8	47.4	142.3
145	33.0	15.1	45.3



Abundance

Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.90 (144.60 to 145.60): V

