

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071619\
 Data File : VX010914.D
 Acq On : 16 Jul 2019 18:15
 Operator : JC/SP
 Sample : K3791-11DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampled :
 FL-0602DL

Quant Time: Jul 17 05:23:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 17 04:37:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	212163	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	331795	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	311342	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	155062	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.06	65	110094	48.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.56%	
35) Dibromofluoromethane	5.49	113	101290	48.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.20%	
50) Toluene-d8	8.71	98	399729	50.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.22%	
62) 4-Bromofluorobenzene	11.13	95	143746	49.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.02%	

Target Compounds

						Qvalue
31) Cyclohexane	5.57	56	56758	18.520	ug/l	98
39) Methylcyclohexane	7.46	83	79373	22.458	ug/l	99
67) Ethyl Benzene	10.25	91	20747	1.906	ug/l	99
68) m/p-Xylenes	10.36	106	34052	8.156	ug/l	96
69) o-Xylene	10.69	106	7468	1.833	ug/l	85
73) Isopropylbenzene	11.02	105	67819	6.473	ug/l	100
78) n-propylbenzene	11.36	91	106892	9.233	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	173682	19.838	ug/l	99
83) tert-Butylbenzene	11.77	119	3300	0.384	ug/l	75
84) 1,2,4-Trimethylbenzene	11.80	105	499217	56.574	ug/l	99
85) sec-Butylbenzene	11.94	105	24269	2.408	ug/l	81
86) p-Isopropyltoluene	12.06	119	19584	2.102	ug/l	95
89) n-Butylbenzene	12.38	91	17946	2.252	ug/l #	18
95) Naphthalene	13.83	128	60216	6.053	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071619\
 Data File : VX010914.D
 Acq On : 16 Jul 2019 18:15
 Operator : JC/SP
 Sample : K3791-11DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampled :
 FL-0602DL

Quant Time: Jul 17 05:23:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 17 04:37:55 2019
 Response via : Initial Calibration

