

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071619\
 Data File : VX010905.D
 Acq On : 16 Jul 2019 13:33
 Operator : JC/SP
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX071619

Quant Time: Jul 17 04:48:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 17 04:37:55 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	107	0.00
2 T	Dichlorodifluoromethane	0.467	0.550	-17.8	112	0.00
3 P	Chloromethane	0.460	0.460	0.0	108	0.00
4 C	Vinyl Chloride	0.469	0.484	-3.2#	107	0.00
5 T	Bromomethane	0.264	0.262	0.8	114	0.00
6 T	Chloroethane	0.298	0.296	0.7	107	0.00
7 T	Trichlorofluoromethane	0.794	0.808	-1.8	109	0.00
8 T	Diethyl Ether	0.274	0.261	4.7	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.453	0.466	-2.9	113	0.00
10 T	Methyl Iodide	0.595	0.563	5.4	96	0.00
11 T	Tert butyl alcohol	0.124	0.109	12.1	102	0.00
12 CM	1,1-Dichloroethene	0.445	0.442	0.7#	106	0.00
13 T	Acrolein	0.072	0.066	8.3	103	0.00
14 T	Allyl chloride	0.647	0.655	-1.2	110	0.00
15 T	Acrylonitrile	0.250	0.232	7.2	102	0.00
16 T	Acetone	0.240	0.233	2.9	106	0.00
17 T	Carbon Disulfide	1.142	1.189	-4.1	111	0.00
18 T	Methyl Acetate	0.547	0.480	12.2	103	0.00
19 T	Methyl tert-butyl Ether	1.434	1.413	1.5	110	0.00
20 T	Methylene Chloride	0.509	0.480	5.7	106	0.00
21 T	trans-1,2-Dichloroethene	0.473	0.472	0.2	109	0.00
22 T	Diisopropyl ether	1.364	1.317	3.4	107	0.00
23 T	Vinyl Acetate	1.156	1.176	-1.7	107	0.00
24 P	1,1-Dichloroethane	0.792	0.767	3.2	108	0.00
25 T	2-Butanone	0.345	0.335	2.9	105	0.00
26 T	2,2-Dichloropropane	0.651	0.685	-5.2	115	0.00
27 T	cis-1,2-Dichloroethene	0.547	0.534	2.4	110	0.00
28 T	Bromochloromethane	0.373	0.346	7.2	105	0.00
29 T	Tetrahydrofuran	0.218	0.204	6.4	102	0.00
30 C	Chloroform	0.856	0.828	3.3#	109	0.00
31 T	Cyclohexane	0.722	0.727	-0.7	112	0.00
32 T	1,1,1-Trichloroethane	0.741	0.744	-0.4	109	0.00
33 S	1,2-Dichloroethane-d4	0.532	0.499	6.2	105	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
35 S	Dibromofluoromethane	0.317	0.307	3.2	103	0.00
36 T	1,1-Dichloropropene	0.411	0.425	-3.4	111	0.00
37 T	Ethyl Acetate	0.421	0.415	1.4	102	0.00
38 T	Carbon Tetrachloride	0.428	0.457	-6.8	113	0.00
39 T	Methylcyclohexane	0.533	0.565	-6.0	114	0.00
40 TM	Benzene	1.258	1.261	-0.2	108	0.00
41 T	Methacrylonitrile	0.239	0.221	7.5	105	0.00
42 TM	1,2-Dichloroethane	0.437	0.442	-1.1	112	0.00
43 T	Isopropyl Acetate	0.692	0.685	1.0	106	0.00
44 TM	Trichloroethene	0.372	0.374	-0.5	109	0.00
45 C	1,2-Dichloropropane	0.320	0.319	0.3#	108	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX071619\
 Data File : VX010905.D
 Acq On : 16 Jul 2019 13:33
 Operator : JC/SP
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX071619

Quant Time: Jul 17 04:48:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 17 04:37:55 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.229	0.230	-0.4	107	0.00
47 T	Bromodichloromethane	0.407	0.437	-7.4	112	0.00
48 T	Methyl methacrylate	0.337	0.339	-0.6	108	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	104	0.00
50 S	Toluene-d8	1.190	1.150	3.4	102	0.00
51 T	4-Methyl-2-Pentanone	0.448	0.442	1.3	104	0.00
52 CM	Toluene	0.826	0.831	-0.6#	108	0.00
53 T	t-1,3-Dichloropropene	0.438	0.479	-9.4	111	0.00
54 T	cis-1,3-Dichloropropene	0.487	0.523	-7.4	109	0.00
55 T	1,1,2-Trichloroethane	0.340	0.336	1.2	106	0.00
56 T	Ethyl methacrylate	0.498	0.520	-4.4	107	0.00
57 T	1,3-Dichloropropane	0.540	0.535	0.9	107	0.00
58 T	2-Chloroethyl Vinyl ether	0.238	0.270	-13.4	117	0.00
59 T	2-Hexanone	0.348	0.357	-2.6	105	0.00
60 T	Dibromochloromethane	0.345	0.375	-8.7	109	0.00
61 T	1,2-Dibromoethane	0.361	0.372	-3.0	108	0.00
62 S	4-Bromofluorobenzene	0.438	0.446	-1.8	106	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
64 T	Tetrachloroethene	0.393	0.390	0.8	110	0.00
65 PM	Chlorobenzene	1.027	1.018	0.9	111	0.00
66 T	1,1,1,2-Tetrachloroethane	0.361	0.376	-4.2	109	0.00
67 C	Ethyl Benzene	1.748	1.766	-1.0#	109	0.00
68 T	m/p-Xylenes	0.671	0.692	-3.1	111	0.00
69 T	o-Xylene	0.654	0.661	-1.1	108	0.00
70 T	Styrene	1.091	1.148	-5.2	109	0.00
71 P	Bromoform	0.288	0.315	-9.4	111	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
73 T	Isopropylbenzene	3.378	3.388	-0.3	112	0.00
74 T	N-amyl acetate	1.291	1.319	-2.2	108	0.00
75 P	1,1,2,2-Tetrachloroethane	1.106	1.071	3.2	107	0.00
76 T	1,2,3-Trichloropropane	0.902	0.897	0.6	108	0.00
77 T	Bromobenzene	0.932	0.902	3.2	109	0.00
78 T	n-propylbenzene	3.733	3.850	-3.1	113	0.00
79 T	2-Chlorotoluene	2.242	2.271	-1.3	112	0.00
80 T	1,3,5-Trimethylbenzene	2.823	2.862	-1.4	110	0.00
81 T	trans-1,4-Dichloro-2-butene	0.328	0.337	-2.7	110	0.00
82 T	4-Chlorotoluene	2.610	2.662	-2.0	113	0.00
83 T	tert-Butylbenzene	2.773	2.842	-2.5	112	0.00
84 T	1,2,4-Trimethylbenzene	2.845	2.912	-2.4	111	0.00
85 T	sec-Butylbenzene	3.249	3.384	-4.2	112	0.00
86 T	p-Isopropyltoluene	3.004	3.185	-6.0	114	0.00
87 T	1,3-Dichlorobenzene	1.617	1.647	-1.9	113	0.00
88 T	1,4-Dichlorobenzene	1.659	1.661	-0.1	112	0.00
89 T	n-Butylbenzene	2.570	2.845	-10.7	117	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071619\
Data File : VX010905.D
Acq On : 16 Jul 2019 13:33
Operator : JC/SP
Sample : VSTDICV050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX071619

Quant Time: Jul 17 04:48:29 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 17 04:37:55 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.474	0.507	-7.0	114	0.00
91 T	1,2-Dichlorobenzene	1.603	1.621	-1.1	111	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.231	0.234	-1.3	108	0.00
93 T	1,2,4-Trichlorobenzene	1.086	1.157	-6.5	115	0.00
94 T	Hexachlorobutadiene	0.538	0.582	-8.2	119	0.00
95 T	Naphthalene	3.208	3.376	-5.2	111	0.00
96 T	1,2,3-Trichlorobenzene	1.081	1.122	-3.8	112	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6