

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071620\
 Data File : VX017395.D
 Acq On : 16 Jul 2020 16:51
 Operator : JC/SP
 Sample : PB130232ZHE#01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130232ZHE#01

Quant Time: Jul 17 01:59:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 14 02:59:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	264622	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	466827	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	529387	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	272885	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	180422	54.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.64%	
35) Dibromofluoromethane	5.47	113	155675	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	
50) Toluene-d8	8.70	98	607080	53.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.04%	
62) 4-Bromofluorobenzene	11.12	95	249705	55.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.84%	
Target Compounds						
16) Acetone	2.42	43	32673	27.575	ug/l	96
18) Methyl Acetate	2.75	43	3984	1.361	ug/l	92
20) Methylene Chloride	2.84	84	26978	9.332	ug/l	98
43) Isopropyl Acetate	6.42	43	92359	13.016	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071620\
Data File : VX017395.D
Acq On : 16 Jul 2020 16:51
Operator : JC/SP
Sample : PB130232ZHE#01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130232ZHE#01

Quant Time: Jul 17 01:59:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 14 02:59:51 2020
Response via : Initial Calibration

