

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071620\
 Data File : VX017396.D
 Acq On : 16 Jul 2020 17:14
 Operator : JC/SP
 Sample : PB130232ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130232ZHE#02

Quant Time: Jul 17 02:01:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 14 02:59:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	265371	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	462025	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	511509	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	254179	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	181618	54.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.06%	
35) Dibromofluoromethane	5.47	113	158236	50.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.68%	
50) Toluene-d8	8.70	98	604139	53.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.62%	
62) 4-Bromofluorobenzene	11.12	95	240531	53.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.88%	
Target Compounds						
16) Acetone	2.42	43	31270	26.317	ug/l	95
18) Methyl Acetate	2.75	43	3541	1.207	ug/l #	80
20) Methylene Chloride	2.84	84	25402	8.762	ug/l	90
43) Isopropyl Acetate	6.41	43	91281	12.998	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071620\
Data File : VX017396.D
Acq On : 16 Jul 2020 17:14
Operator : JC/SP
Sample : PB130232ZHE#02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130232ZHE#02

Quant Time: Jul 17 02:01:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 14 02:59:51 2020
Response via : Initial Calibration

