

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071620\
 Data File : VX017402.D
 Acq On : 16 Jul 2020 19:30
 Operator : JC/SP
 Sample : L3324-05 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOTE-50-DIAZ

Quant Time: Jul 17 02:25:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 14 02:59:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	246934	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	435369	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	469877	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	255339	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	172361	55.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.22%	
35) Dibromofluoromethane	5.46	113	151590	51.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.38%	
50) Toluene-d8	8.70	98	549549	51.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.90%	
62) 4-Bromofluorobenzene	11.12	95	233141	55.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.96%	
Target Compounds						
16) Acetone	2.42	43	13663	12.357	ug/l	97
20) Methylene Chloride	2.83	84	4295	1.592	ug/l #	85
25) 2-Butanone	4.65	43	4933	2.809	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071620\
Data File : VX017402.D
Acq On : 16 Jul 2020 19:30
Operator : JC/SP
Sample : L3324-05 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOTE-50-DIAZ

Quant Time: Jul 17 02:25:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 14 02:59:51 2020
Response via : Initial Calibration

