

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071620\
 Data File : VX017406.D
 Acq On : 16 Jul 2020 21:01
 Operator : JC/SP
 Sample : L3334-03 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 I-LDC-4

Quant Time: Jul 17 02:46:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 14 02:59:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	246600	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	432580	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	458275	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	242622	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	171848	55.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.04%	
35) Dibromofluoromethane	5.46	113	152003	52.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.32%	
50) Toluene-d8	8.70	98	538147	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
62) 4-Bromofluorobenzene	11.12	95	219596	52.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.20%	
Target Compounds						
					Qvalue	
16) Acetone	2.42	43	16971	15.370	ug/l	99
17) Carbon Disulfide	2.55	76	358632	54.782	ug/l	100
20) Methylene Chloride	2.84	84	2907	1.079	ug/l #	82
30) Chloroform	5.18	83	11404	2.435	ug/l	88
40) Benzene	6.12	78	9686	0.847	ug/l #	89
52) Toluene	8.77	92	2513	0.342	ug/l	91
68) m/p-Xylenes	10.34	106	2318	0.393	ug/l	91
95) Naphthalene	13.82	128	447527	30.110	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071620\
Data File : VX017406.D
Acq On : 16 Jul 2020 21:01
Operator : JC/SP
Sample : L3334-03 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
I-LDC-4

Quant Time: Jul 17 02:46:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 14 02:59:51 2020
Response via : Initial Calibration

