

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071620\
 Data File : VX017408.D
 Acq On : 16 Jul 2020 21:47
 Operator : JC/SP
 Sample : L3328-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DAY-5

Quant Time: Jul 17 02:55:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 14 02:59:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	248779	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	442892	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	459943	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	251292	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	176519	56.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.06%	
35) Dibromofluoromethane	5.47	113	149903	50.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.50%	
50) Toluene-d8	8.70	98	548314	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
62) 4-Bromofluorobenzene	11.12	95	226682	53.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.06%	
Target Compounds						
					Qvalue	
16) Acetone	2.42	43	6600	5.925	ug/l	99
17) Carbon Disulfide	2.55	76	7050	1.067	ug/l #	94
20) Methylene Chloride	2.84	84	2973	1.094	ug/l #	95
31) Cyclohexane	5.54	56	1781	0.495	ug/l #	76
39) Methylcyclohexane	7.45	83	4186	0.946	ug/l #	86
85) sec-Butylbenzene	11.93	105	5709	0.377	ug/l	99
89) n-Butylbenzene	12.37	91	3731	0.306	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071620\
Data File : VX017408.D
Acq On : 16 Jul 2020 21:47
Operator : JC/SP
Sample : L3328-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DAY-5

Quant Time: Jul 17 02:55:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 14 02:59:51 2020
Response via : Initial Calibration

