


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX071620\
 Data File : VX017383.D
 Acq On : 16 Jul 2020 11:06
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Title : SW846 8260
 Last Update : Tue Jul 14 02:59:51 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1636

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.3	11737	PASS
75	95	30	60	49.3	31618	PASS
95	95	100	100	100.0	64160	PASS
96	95	5	9	6.8	4368	PASS
173	174	0.00	2	1.5	801	PASS
174	95	50	100	83.3	53440	PASS
175	174	5	9	7.2	3832	PASS
176	174	95	101	97.8	52282	PASS
177	176	5	9	6.9	3622	PASS