

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071621\
 Data File : VX023315.D
 Acq On : 16 Jul 2021 09:46
 Operator : JC/MD
 Sample : VX0716MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0716MBL01

Quant Time: Jul 17 02:06:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	200762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	326054	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	277459	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	109337	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	137406	50.924	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.840%
35) Dibromofluoromethane	5.391	113	105423	50.070	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.140%
50) Toluene-d8	8.653	98	380948	48.503	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.000%
62) 4-Bromofluorobenzene	11.085	95	130417	43.012	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071621\
Data File : VX023315.D
Acq On : 16 Jul 2021 09:46
Operator : JC/MD
Sample : VX0716MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0716MBL01

Quant Time: Jul 17 02:06:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

