

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071621\
 Data File : VX023316.D
 Acq On : 16 Jul 2021 10:20
 Operator : JC/MD
 Sample : VX0716WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0716WBL01

Quant Time: Jul 17 02:06:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	229369	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	372181	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	330215	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	140673	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	157588	51.120	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.240%
35) Dibromofluoromethane	5.391	113	122246	50.864	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.720%
50) Toluene-d8	8.653	98	438754	48.940	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.880%
62) 4-Bromofluorobenzene	11.085	95	158426	45.774	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.540%
Target Compounds						
16) Acetone	2.386	43	5669	4.648	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071621\
Data File : VX023316.D
Acq On : 16 Jul 2021 10:20
Operator : JC/MD
Sample : VX0716WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0716WBL01

Quant Time: Jul 17 02:06:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

