

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071621\
 Data File : VX023319.D
 Acq On : 16 Jul 2021 11:46
 Operator : JC/MD
 Sample : M3061-02DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB-2021-WW-000-08DL

Quant Time: Jul 17 02:08:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	219803	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	358419	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	321863	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	139059	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	151363	51.237	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.480%
35) Dibromofluoromethane	5.397	113	117726	50.864	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.720%
50) Toluene-d8	8.653	98	427424	49.506	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.020%
62) 4-Bromofluorobenzene	11.085	95	156571	46.975	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.940%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	6153	5.265	ug/l	89
17) Carbon Disulfide	2.514	76	1596	0.386	ug/l #	90
19) Methyl tert-butyl Ether	3.123	73	28783	3.928	ug/l #	84
44) Trichloroethene	7.129	130	2456	0.933	ug/l	97
64) Tetrachloroethene	9.281	164	167748	65.344	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071621\
Data File : VX023319.D
Acq On : 16 Jul 2021 11:46
Operator : JC/MD
Sample : M3061-02DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB-2021-WW-000-08DL

Quant Time: Jul 17 02:08:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

