

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071621\
 Data File : VX023320.D
 Acq On : 16 Jul 2021 12:21
 Operator : JC/MD
 Sample : M3027-03 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NWB-1639

Quant Time: Jul 17 02:08:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	236448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	386471	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	366767	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	178179	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	168022	52.873	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 105.740%		
35) Dibromofluoromethane	5.385	113	125440	50.263	ug/l	-0.01
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.520%		
50) Toluene-d8	8.653	98	463021	49.737	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.480%		
62) 4-Bromofluorobenzene	11.085	95	190625	53.040	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 106.080%		
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	56673	45.077	ug/l	93
18) Methyl Acetate	2.727	43	3132	1.026	ug/l #	85
20) Methylene Chloride	2.782	84	4466	1.650	ug/l	93
95) Naphthalene	13.780	128	9818	1.995	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071621\
Data File : VX023320.D
Acq On : 16 Jul 2021 12:21
Operator : JC/MD
Sample : M3027-03 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NWB-1639

Quant Time: Jul 17 02:08:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

