

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071621\
 Data File : VX023314.D
 Acq On : 16 Jul 2021 09:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 17 02:05:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.453	0.474	-4.6	92	0.00
3 P	Chloromethane	0.448	0.387	13.6	82	0.00
4 C	Vinyl Chloride	0.475	0.437	8.0#	84	0.00
5 T	Bromomethane	0.265	0.205	22.6	78	0.00
6 T	Chloroethane	0.310	0.274	11.6	81	0.00
7 T	Trichlorofluoromethane	0.825	0.795	3.6	87	0.00
8 T	Diethyl Ether	0.305	0.285	6.6	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.483	0.510	-5.6	98	0.00
10 T	Methyl Iodide	0.593	0.571	3.7	90	0.00
11 T	Tert butyl alcohol	0.141	0.130	7.8	92	0.00
12 CM	1,1-Dichloroethene	0.446	0.441	1.1#	91	0.00
13 T	Acrolein	0.094	0.082	12.8	83	0.00
14 T	Allyl chloride	0.774	0.791	-2.2	94	0.00
15 T	Acrylonitrile	0.275	0.276	-0.4	93	0.00
16 T	Acetone	0.266	0.245	7.9	91	0.00
17 T	Carbon Disulfide	0.940	0.873	7.1	89	0.00
18 T	Methyl Acetate	0.645	0.683	-5.9	99	0.00
19 T	Methyl tert-butyl Ether	1.667	1.775	-6.5	98	0.00
20 T	Methylene Chloride	0.572	0.513	10.3	91	0.00
21 T	trans-1,2-Dichloroethene	0.481	0.464	3.5	91	0.00
22 T	Diisopropyl ether	1.633	1.626	0.4	92	0.00
23 T	Vinyl Acetate	1.382	1.403	-1.5	91	0.00
24 P	1,1-Dichloroethane	0.923	0.934	-1.2	94	0.00
25 T	2-Butanone	0.401	0.400	0.2	92	0.00
26 T	2,2-Dichloropropane	0.774	0.903	-16.7	107	0.00
27 T	cis-1,2-Dichloroethene	0.583	0.581	0.3	93	0.00
28 T	Bromochloromethane	0.438	0.420	4.1	92	0.00
29 T	Tetrahydrofuran	0.258	0.250	3.1	90	0.00
30 C	Chloroform	0.997	1.022	-2.5#	95	0.00
31 T	Cyclohexane	0.805	0.736	8.6	88	0.00
32 T	1,1,1-Trichloroethane	0.868	0.954	-9.9	101	0.00
33 S	1,2-Dichloroethane-d4	0.672	0.652	3.0	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.323	0.325	-0.6	97	-0.01
36 T	1,1-Dichloropropene	0.446	0.448	-0.4	92	0.00
37 T	Ethyl Acetate	0.486	0.502	-3.3	92	0.00
38 T	Carbon Tetrachloride	0.465	0.539	-15.9	102	0.00
39 T	Methylcyclohexane	0.521	0.534	-2.5	94	0.00
40 TM	Benzene	1.294	1.303	-0.7	91	0.00
41 T	Methacrylonitrile	0.272	0.284	-4.4	94	0.00
42 TM	1,2-Dichloroethane	0.520	0.551	-6.0	93	0.00
43 T	Isopropyl Acetate	0.809	0.842	-4.1	93	0.00
44 TM	Trichloroethene	0.367	0.381	-3.8	94	0.00
45 C	1,2-Dichloropropane	0.338	0.343	-1.5#	92	0.00
46 T	Dibromomethane	0.241	0.252	-4.6	93	0.00
47 T	Bromodichloromethane	0.441	0.508	-15.2	100	0.00
48 T	Methyl methacrylate	0.397	0.413	-4.0	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071621\
 Data File : VX023314.D
 Acq On : 16 Jul 2021 09:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 17 02:05:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	80	0.00
50 S	Toluene-d8	1.204	1.160	3.7	91	0.00
51 T	4-Methyl-2-Pentanone	0.501	0.521	-4.0	91	0.00
52 CM	Toluene	0.838	0.859	-2.5#	91	0.00
53 T	t-1,3-Dichloropropene	0.481	0.564	-17.3	99	0.00
54 T	cis-1,3-Dichloropropene	0.510	0.591	-15.9	97	0.00
55 T	1,1,2-Trichloroethane	0.352	0.378	-7.4	96	0.00
56 T	Ethyl methacrylate	0.543	0.586	-7.9	95	0.00
57 T	1,3-Dichloropropane	0.581	0.601	-3.4	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.266	0.293	-10.2	91	0.00
59 T	2-Hexanone	0.385	0.421	-9.4	95	0.00
60 T	Dibromochloromethane	0.338	0.405	-19.8	102	0.00
61 T	1,2-Dibromoethane	0.361	0.392	-8.6	93	0.00
62 S	4-Bromofluorobenzene	0.465	0.477	-2.6	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.399	0.425	-6.5	96	0.00
65 PM	Chlorobenzene	0.988	1.038	-5.1	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.357	0.418	-17.1	99	0.00
67 C	Ethyl Benzene	1.747	1.838	-5.2#	92	0.00
68 T	m/p-Xylenes	0.667	0.724	-8.5	94	0.00
69 T	o-Xylene	0.672	0.715	-6.4	93	0.00
70 T	Styrene	1.098	1.219	-11.0	94	0.00
71 P	Bromoform	0.252	0.321	-27.4#	106	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.522	3.705	-5.2	94	0.00
74 T	N-amyl acetate	1.495	1.498	-0.2	89	0.00
75 P	1,1,2,2-Tetrachloroethane	1.119	1.121	-0.2	93	0.00
76 T	1,2,3-Trichloropropane	1.006	1.116	-10.9	98	0.00
77 T	Bromobenzene	0.879	0.902	-2.6	93	0.00
78 T	n-propylbenzene	4.073	4.264	-4.7	93	0.00
79 T	2-Chlorotoluene	2.517	2.602	-3.4	94	0.00
80 T	1,3,5-Trimethylbenzene	3.033	3.216	-6.0	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.328	0.385	-17.4	103	0.00
82 T	4-Chlorotoluene	2.843	3.027	-6.5	94	0.00
83 T	tert-Butylbenzene	2.879	3.131	-8.8	96	0.00
84 T	1,2,4-Trimethylbenzene	2.997	3.210	-7.1	94	0.00
85 T	sec-Butylbenzene	3.704	4.027	-8.7	96	0.00
86 T	p-Isopropyltoluene	3.107	3.458	-11.3	97	0.00
87 T	1,3-Dichlorobenzene	1.651	1.747	-5.8	95	0.00
88 T	1,4-Dichlorobenzene	1.663	1.759	-5.8	96	0.00
89 T	n-Butylbenzene	2.722	2.993	-10.0	96	0.00
90 T	Hexachloroethane	0.455	0.567	-24.6	106	0.00
91 T	1,2-Dichlorobenzene	1.643	1.727	-5.1	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.238	0.293	-23.1	104	0.00
93 T	1,2,4-Trichlorobenzene	1.012	1.137	-12.4	101	0.00
94 T	Hexachlorobutadiene	0.428	0.519	-21.3	110	0.00
95 T	Naphthalene	3.366	3.823	-13.6	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071621\
Data File : VX023314.D
Acq On : 16 Jul 2021 09:13
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 17 02:05:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.018	1.165	-14.4	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6