

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071625\
 Data File : VX047029.D
 Acq On : 16 Jul 2025 14:46
 Operator : JC/MD
 Sample : Q2605-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 V908

Quant Time: Jul 17 03:35:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	448380	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	815427	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	763944	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	386056	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	326188	55.066	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 110.140%			
35) Dibromofluoromethane	5.397	113	267433	48.316	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.640%			
50) Toluene-d8	8.653	98	979053	50.133	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.260%			
62) 4-Bromofluorobenzene	11.079	95	404599	54.512	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 109.020%			
Target Compounds						
16) Acetone	2.392	43	39485	21.460	ug/l	99
18) Methyl Acetate	2.715	43	35608	7.342	ug/l	96
20) Methylene Chloride	2.806	84	18458	3.392	ug/l	87
37) Ethyl Acetate	4.751	43	16077m	2.528	ug/l	
43) Isopropyl Acetate	6.367	43	10952	1.124	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071625\
Data File : VX047029.D
Acq On : 16 Jul 2025 14:46
Operator : JC/MD
Sample : Q2605-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
V908

Quant Time: Jul 17 03:35:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

