

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071625\
 Data File : VX047030.D
 Acq On : 16 Jul 2025 15:07
 Operator : JC/MD
 Sample : Q2605-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VB16135

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 03:35:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	412937	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	696709	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	643987	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	333810	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	275859	50.567	ug/l	0.01
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.140%			
35) Dibromofluoromethane	5.397	113	223164	47.188	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 94.380%			
50) Toluene-d8	8.653	98	832298	49.880	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 99.760%			
62) 4-Bromofluorobenzene	11.079	95	339814	53.585	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 107.160%			
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	39433	23.271	ug/l	95
18) Methyl Acetate	2.721	43	36618	8.199	ug/l	98
20) Methylene Chloride	2.806	84	18368	3.666	ug/l	92
37) Ethyl Acetate	4.751	43	21620m	3.978	ug/l	
43) Isopropyl Acetate	6.361	43	14549	1.748	ug/l #	92
52) Toluene	8.726	92	16248	1.359	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071625\
Data File : VX047030.D
Acq On : 16 Jul 2025 15:07
Operator : JC/MD
Sample : Q2605-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VB16135

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 03:35:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

