

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071625\
 Data File : VX047032.D
 Acq On : 16 Jul 2025 15:51
 Operator : JC/MD
 Sample : Q2605-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 V897

Quant Time: Jul 17 03:36:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	441027	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	774825	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	726032	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	377203	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	311717	53.501	ug/l	0.01
Spiked Amount 50.000	Range 78 - 117		Recovery = 107.000%			
35) Dibromofluoromethane	5.404	113	254672	48.421	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.840%			
50) Toluene-d8	8.653	98	931198	50.181	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.360%			
62) 4-Bromofluorobenzene	11.079	95	386572	54.812	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 109.620%			
Target Compounds					Qvalue	
16) Acetone	2.392	43	41919	23.162	ug/l	100
18) Methyl Acetate	2.721	43	36405	7.632	ug/l	97
20) Methylene Chloride	2.806	84	18736	3.501	ug/l	96
37) Ethyl Acetate	4.745	43	20315m	3.361	ug/l	
43) Isopropyl Acetate	6.367	43	15385	1.662	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071625\
Data File : VX047032.D
Acq On : 16 Jul 2025 15:51
Operator : JC/MD
Sample : Q2605-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
V897

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 03:36:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

