

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071718\
 Data File : VX003426.D
 Acq On : 17 Jul 2018 14:33
 Operator : JC/MD
 Sample : J4006-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RR-BASELINE-WATER

Quant Time: Jul 18 01:12:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071618W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 16 13:15:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	294205	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	428447	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	388160	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	240060	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	176565	45.83	ug/l	0.00
Spiked Amount			Recovery	=	91.66%	
35) Dibromofluoromethane	5.51	113	161343	46.54	ug/l	0.00
Spiked Amount			Recovery	=	93.08%	
50) Toluene-d8	8.72	98	694746	54.31	ug/l	0.00
Spiked Amount			Recovery	=	108.62%	
62) 4-Bromofluorobenzene	11.14	95	225781	49.43	ug/l	0.00
Spiked Amount			Recovery	=	98.86%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	4085926	2549.64	ug/l	# 87
17) Carbon Disulfide	2.57	76	17345	2.04	ug/l	96
25) 2-Butanone	4.72	43	1217029	485.26	ug/l	99
29) Tetrahydrofuran	5.17	42	1197517	752.90	ug/l	91
40) Benzene	6.15	78	18926	1.43	ug/l	97
51) 4-Methyl-2-Pentanone	8.66	43	5526	1.14	ug/l	93
52) Toluene	8.79	92	28866	3.49	ug/l	99
67) Ethyl Benzene	10.26	91	8756	0.58	ug/l	100
68) m/p-Xylenes	10.36	106	13089	2.26	ug/l	99
69) o-Xylene	10.70	106	6729	1.21	ug/l	98
70) Styrene	10.72	104	4685	0.51	ug/l	# 87
80) 1,3,5-Trimethylbenzene	11.51	105	3822	0.30	ug/l	96
84) 1,2,4-Trimethylbenzene	11.81	105	10359	0.80	ug/l	100
95) Naphthalene	13.83	128	58747	3.73	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071718\
Data File : VX003426.D
Acq On : 17 Jul 2018 14:33
Operator : JC/MD
Sample : J4006-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RR-BASELINE-WATER

Quant Time: Jul 18 01:12:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071618W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 16 13:15:28 2018
Response via : Initial Calibration

